

Data File : VW011554.D
Acq On : 29 Jul 2019 12:35
Operator : SY/VA
Sample : K3979-17
Misc : 4.54G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DB936

Quant Time: Jul 30 05:59:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072219S.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 30 05:56:36 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	471609	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	420311	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	192105	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	124848	19.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.32%
7) Chloroethane-d5	2.89	69	96969	21.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.56%
10) 1,1-Dichloroethene-d2	4.02	63	221440	15.52	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.08%
20) 2-Butanone-d5	7.07	46	97192	36.25	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.50%
24) Chloroform-d	7.65	84	276029	22.86	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.44%
26) 1,2-Dichloroethane-d4	8.30	65	160891	22.41	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.64%
29) Benzene-d6	8.27	84	570214	22.68	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.72%
33) 1,2-Dichloropropane-d6	9.27	67	186655	22.92	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.68%
37) Toluene-d8	10.32	98	499701	22.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.72%
38) trans-1,3-Dichloropropene-	10.58	79	74906	21.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.16%
39) 2-Hexanone-d5	10.93	63	65576	37.14	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.28%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	131581	20.22	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.88%
61) 1,2-Dichlorobenzene-d4	13.85	152	168142	23.31	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.24%


Target Compounds					Qvalue
13) Acetone	4.12	43	3838	1.353 ug/L	82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW011554.D
Acq On : 29 Jul 2019 12:35
Operator : SY/VA
Sample : K3979-17
Misc : 4.54G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DB936

Quant Time: Jul 30 05:59:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072219S.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 30 05:56:36 2019
Response via : Initial Calibration





#13

Acetone

Concen: 1.353 ug/L

RT: 4.12 min Scan# 364

Delta R.T. -0.00 min

Lab File: VW011554.D

Acq: 29 Jul 2019 12:35

Instrument :

MSVOA_W

ClientSampleId :

DB936

Tgt Ion: 43 Resp: 3838

Ion Ratio Lower Upper

43 100

58 19.3 0.0 57.2

