

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW073020\
 Data File : VW016012.D
 Acq On : 30 Jul 2020 15:32
 Operator : SY/VA
 Sample : L3492-18
 Misc : 3.61G/10ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 D0AB4

Quant Time: Jul 31 07:21:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 31 06:34:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	422271	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	395705	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	163312	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	112205	18.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.32%
7) Chloroethane-d5	2.89	69	93382	19.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.20%
10) 1,1-Dichloroethene-d2	4.01	63	174471	14.48	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.92%
20) 2-Butanone-d5	7.09	46	70868	45.28	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.56%
24) Chloroform-d	7.65	84	244876	20.65	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.60%
26) 1,2-Dichloroethane-d4	8.31	65	136005	20.89	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.56%
29) Benzene-d6	8.28	84	485540	20.16	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.64%
33) 1,2-Dichloropropane-d6	9.28	67	151759	21.02	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.08%
37) Toluene-d8	10.33	98	422886	19.38	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.52%
38) trans-1,3-Dichloropropene-	10.58	79	58636	19.05	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.20%
39) 2-Hexanone-d5	10.93	63	51244	42.62	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.24%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	123744	22.29	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.16%
61) 1,2-Dichlorobenzene-d4	13.86	152	134217	21.08	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.32%

Target Compounds					Ovalue
13) Acetone	4.14	43	4843	3.684 ug/L	96
16) Methylene chloride	4.92	84	32213	4.655 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW073020\
Data File : VW016012.D
Acq On : 30 Jul 2020 15:32
Operator : SY/VA
Sample : L3492-18
Misc : 3.61G/10ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
D0AB4

Quant Time: Jul 31 07:21:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 31 06:34:21 2020
Response via : Initial Calibration

