

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW073020\
 Data File : VW016014.D
 Acq On : 30 Jul 2020 16:20
 Operator : SY/VA
 Sample : L3492-20
 Misc : 4.05G/10ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 D0AB6

Quant Time: Jul 31 07:25:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 31 06:34:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	405888	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	353950	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	121632	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	113567	19.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.32%
7) Chloroethane-d5	2.90	69	93485	20.10	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.40%
10) 1,1-Dichloroethene-d2	4.01	63	174302	15.05	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.20%
20) 2-Butanone-d5	7.09	46	69272	46.04	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.08%
24) Chloroform-d	7.65	84	241752	21.21	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.84%
26) 1,2-Dichloroethane-d4	8.31	65	133095	21.27	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.08%
29) Benzene-d6	8.28	84	479261	22.25	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.00%
33) 1,2-Dichloropropane-d6	9.28	67	152889	23.67	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.68%
37) Toluene-d8	10.33	98	412092	21.12	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.48%
38) trans-1,3-Dichloropropene-	10.59	79	58424	21.22	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.88%
39) 2-Hexanone-d5	10.93	63	49131	45.68	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.36%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	114412	23.05	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.20%
61) 1,2-Dichlorobenzene-d4	13.86	152	106898	22.54	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.16%


Target Compounds					Ovalue
13) Acetone	4.14	43	3791	3.001 ug/L	85
16) Methylene chloride	4.92	84	31448	4.728 ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW073020\
Data File : VW016014.D
Acq On : 30 Jul 2020 16:20
Operator : SY/VA
Sample : L3492-20
Misc : 4.05G/10ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
D0AB6

Quant Time: Jul 31 07:25:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 31 06:34:21 2020
Response via : Initial Calibration





#13

Acetone

Concen: 3.001 ug/L

RT: 4.14 min Scan# 367

Delta R.T. -0.00 min

Lab File: VW016014.D

Acq: 30 Jul 2020 16:20

Instrument :

MSVOA_W

ClientSampled :





D0AB6

Tot Ion: 43 Resp: 3791

Ion Ratio Lower Upper

43 100

58 20.8 0.0 58.2



#16

Methylene chloride

Concen: 4.728 ug/L

RT: 4.92 min Scan# 495

Delta R.T. 0.01 min

Lab File: VW016014.D

Acq: 30 Jul 2020 16:20

Tgt Ion: 84 Resp: 31448

Ion Ratio Lower Upper

84 100

49 123.7 91.1 169.1

86 63.9 46.0 85.4

