

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW073124\
 Data File : VW029810.D
 Acq On : 31 Jul 2024 12:19
 Operator : SY/MD
 Sample : VIBLK504
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK504

Quant Time: Aug 01 04:44:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Aug 01 04:42:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	251568	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	245251	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	110320	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	60201	16.856	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.440%
7) Chloroethane-d5	2.893	69	49776	19.522	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.080%
11) 1,1-Dichloroethene-d2	4.021	65	31079	19.064	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	76.240%
21) 2-Butanone-d5	7.081	46	42266	43.178	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.360%
24) Chloroform-d	7.648	84	151524	23.268	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.080%
26) 1,2-Dichloroethane-d4	8.301	65	90838	24.490	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.960%
32) Benzene-d6	8.276	84	284248	21.450	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.800%
36) 1,2-Dichloropropane-d6	9.270	67	97549	23.054	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.200%
41) Toluene-d8	10.325	98	256816	21.796	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.200%
43) trans-1,3-Dichloroprop...	10.575	79	36486	21.310	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.240%
47) 2-Hexanone-d5	10.922	63	28237	40.472	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.940%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	73137	23.279	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.120%
66) 1,2-Dichlorobenzene-d4	13.848	152	82778	24.165	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW073124\
Data File : VW029810.D
Acq On : 31 Jul 2024 12:19
Operator : SY/MD
Sample : VIBLK504
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK504

Quant Time: Aug 01 04:44:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Aug 01 04:42:41 2024
Response via : Initial Calibration

