

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080118\
 Data File : VW004395.D
 Acq On : 02 Aug 2018 03:50
 Operator : SY/AP
 Sample : VIBLK86
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK86

Quant Time: Aug 02 05:52:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM073118S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 02 05:11:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	159739	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	159257	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	77147	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	37293	29.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	119.84%
7) Chloroethane-d5	2.89	69	34333	27.93	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	111.72%
10) 1,1-Dichloroethene-d2	4.01	63	59257	16.49	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.96%
20) 2-Butanone-d5	7.09	46	36852	47.75	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.50%
24) Chloroform-d	7.65	84	92004	23.30	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	8.31	65	54975	24.86	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.44%
29) Benzene-d6	8.28	84	170176	20.69	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.76%
33) 1,2-Dichloropropane-d6	9.28	67	48851	19.17	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	76.68%
37) Toluene-d8	10.33	98	163766	20.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.36%
38) trans-1,3-Dichloropropene-	10.58	79	24036	19.52	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.08%
39) 2-Hexanone-d5	10.93	63	30019	45.89	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.78%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	57906	21.76	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.04%
61) 1,2-Dichlorobenzene-d4	13.86	152	72272	24.51	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.04%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW080118\
Data File : VW004395.D
Acq On : 02 Aug 2018 03:50
Operator : SY/AP
Sample : VIBLK86
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK86

Quant Time: Aug 02 05:52:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM073118S.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 02 05:11:26 2018
Response via : Initial Calibration

