


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080319\
Data File : VW011633.D
Acq On : 02 Aug 2019 20:56
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Title : SW846 8260
Last Update : Sat Jul 27 00:23:30 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1747

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.9	26640	PASS
75	95	30	60	55.3	64277	PASS
95	95	100	100	100.0	116293	PASS
96	95	5	9	6.9	8009	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	72.4	84154	PASS
175	174	5	9	8.2	6887	PASS
176	174	95	101	98.9	83269	PASS
177	176	5	9	6.8	5670	PASS