

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080321\
 Data File : VW019663.D
 Acq On : 03 Aug 2021 16:42
 Operator : SY/VA
 Sample : VW0803SBL01
 Misc : 5.00g/5.00mL/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0803SBL01

Quant Time: Aug 04 06:47:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W080321S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 16:22:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.945	168	647407	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.842	114	1102250	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	958455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	463060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	397883	48.306	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	96.620%		
35) Dibromofluoromethane	7.884	113	341895	46.926	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	93.860%		
50) Toluene-d8	10.323	98	1358504	47.713	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	95.420%		
62) 4-Bromofluorobenzene	12.615	95	470149	46.279	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	92.560%		
Target Compounds					Qvalue	
16) Acetone	4.141	43	18634	15.312	ug/l	91
20) Methylene Chloride	4.915	84	132122	11.709	ug/l	97
95) Naphthalene	15.365	128	13498	5.066	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080321\
Data File : VW019663.D
Acq On : 03 Aug 2021 16:42
Operator : SY/VA
Sample : VW0803SBL01
Misc : 5.00g/5.00mL/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0803SBL01

Quant Time: Aug 04 06:47:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W080321S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 16:22:46 2021
Response via : Initial Calibration

