


Methyl Acetate



$$\text{Response} = 2.720\text{e-}001 * \text{Amt} - 6.828\text{e-}003$$


Coef of Det (r^2) = 0.996138 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA W\Method\82W080321S.M

Calibration Table Last Updated: Tue Aug 03 16:22:46 2021

Naphthalene

Response Ratio



$$\text{Response} = 2.059\text{e}+000 * \text{Amt} - 1.794\text{e}-001$$


Coef of Det (r^2) = 0.997931 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA W\Method\82W080321S.M

Calibration Table Last Updated: Tue Aug 03 16:22:46 2021

Methylene Chloride

Response Ratio




$$\text{Response} = 5.381\text{e-}001 * \text{Amt} + 7.807\text{e-}002$$

Coef of Det (r^2) = 0.998504 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA W\Method\82W080321S.M

Calibration Table Last Updated: Tue Aug 03 16:22:46 2021

trans-1,4-Dichloro-2-butene



Response = $2.262 \times 10^{-1} \times \text{Amt} - 1.929 \times 10^{-2}$
Coef of Det (r^2) = 0.997858 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA W\Method\82W080321S.M
Calibration Table Last Updated: Tue Aug 03 16:22:46 2021