

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080419\
 Data File : VW011674.D
 Acq On : 03 Aug 2019 18:48
 Operator : SY/VA
 Sample : K4152-15
 Misc : 5.36G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 2S-E4-(0-1)

Quant Time: Aug 04 02:24:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	214099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	363846	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	296813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	111931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	143303	58.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.58%	
35) Dibromofluoromethane	7.88	113	111870	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
50) Toluene-d8	10.32	98	446690	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
62) 4-Bromofluorobenzene	12.62	95	135189	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
Target Compounds						
16) Acetone	4.13	43	79870	102.132	ug/l	95
20) Methylene Chloride	4.91	84	22980	8.520	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW080419\
Data File : VW011674.D
Acq On : 03 Aug 2019 18:48
Operator : SY/VA
Sample : K4152-15
Misc : 5.36G/5ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
2S-E4-(0-1)

Quant Time: Aug 04 02:24:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:23:30 2019
Response via : Initial Calibration

