

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080522\
 Data File : VW024188.D
 Acq On : 05 Aug 2022 13:32
 Operator : SY/VA
 Sample : N4030-04
 Misc : 5.16g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0QB4

Quant Time: Aug 05 23:45:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 05 23:43:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	51801	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	48234	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	20441	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	33122	30.590	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	122.360%
7) Chloroethane-d5	2.873	69	22681	33.626	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	134.520%
11) 1,1-Dichloroethene-d2	4.013	63	27356	22.169	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	88.680%
21) 2-Butanone-d5	7.080	46	10028	49.781	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.560%
24) Chloroform-d	7.647	84	39495	25.198	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.800%
26) 1,2-Dichloroethane-d4	8.305	65	24687	28.370	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	113.480%
32) Benzene-d6	8.274	84	78020	27.427	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	109.720%
36) 1,2-Dichloropropane-d6	9.274	67	23357	27.809	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	111.240%
41) Toluene-d8	10.323	98	62755	23.602	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.400%
43) trans-1,3-Dichloroprop...	10.579	79	9579	26.039	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	104.160%
47) 2-Hexanone-d5	10.926	63	6050	38.999	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.000%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	21785	26.734	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	106.920%
66) 1,2-Dichlorobenzene-d4	13.853	152	19456	26.010	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.040%
Target Compounds						Qvalue
16) Methylene chloride	4.915	84	5837m	6.390	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080522\
Data File : VW024188.D
Acq On : 05 Aug 2022 13:32
Operator : SY/VA
Sample : N4030-04
Misc : 5.16g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0QB4

Quant Time: Aug 05 23:45:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 05 23:43:03 2022
Response via : Initial Calibration

