

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080522\
 Data File : VW024189.D
 Acq On : 05 Aug 2022 13:56
 Operator : SY/VA
 Sample : N4030-05
 Misc : 4.81g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0QB5

Quant Time: Aug 05 23:46:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 05 23:43:03 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.836	114	429517	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.628	117	394864	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.554	152	183093	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.343	65	242112	26.967	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 107.880%		
7) Chloroethane-d5		2.873	69	133363	23.846	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 95.400%		
11) 1,1-Dichloroethene-d2		4.007	63	204303	19.967	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	= 79.880%		
21) 2-Butanone-d5		7.080	46	79704	47.718	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 95.440%		
24) Chloroform-d		7.647	84	312151	24.019	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 96.080%		
26) 1,2-Dichloroethane-d4		8.305	65	171051	23.707	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 94.840%		
32) Benzene-d6		8.269	84	624830	26.831	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 107.320%		
36) 1,2-Dichloropropane-d6		9.275	67	178454	25.954	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 103.800%		
41) Toluene-d8		10.323	98	568000	26.095	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 104.360%		
43) trans-1,3-Dichloroprop...		10.579	79	75022	24.911	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 99.640%		
47) 2-Hexanone-d5		10.927	63	61219	48.204	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 96.400%		
56) 1,1,2,2-Tetrachloroeth...		12.689	84	139631	20.931	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 83.720%		
66) 1,2-Dichlorobenzene-d4		13.853	152	180169	26.891	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 107.560%		
Target Compounds							
13) Acetone		4.123	43	5358	4.357	ug/L	# 62
16) Methylene chloride		4.916	84	40634	5.365	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080522\
Data File : VW024189.D
Acq On : 05 Aug 2022 13:56
Operator : SY/VA
Sample : N4030-05
Misc : 4.81g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0QB5

Quant Time: Aug 05 23:46:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 05 23:43:03 2022
Response via : Initial Calibration

