

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080619\
 Data File : VW011708.D
 Acq On : 05 Aug 2019 14:27
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 06 08:32:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.279	0.277	0.7	93	0.00
3 P	Chloromethane	0.483	0.441	8.7	91	0.00
4 C	Vinyl Chloride	0.582	0.589	-1.2#	94	0.00
5 T	Bromomethane	0.322	0.307	4.7	91	0.00
6 T	Chloroethane	0.332	0.312	6.0	86	0.00
7 T	Trichlorofluoromethane	0.260	0.316	-21.5#	109	0.00
8 T	Diethyl Ether	0.295	0.290	1.7	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.493	0.538	-9.1	99	-0.01
10 T	Methyl Iodide	0.697	0.697	0.0	89	0.00
11 T	Tert butyl alcohol	0.048	0.034	29.2#	78	0.06
12 CM	1,1-Dichloroethene	0.521	0.544	-4.4#	95	0.00
13 T	Acrolein	0.041	0.040	2.4	96	0.00
14 T	Allyl chloride	1.108	1.173	-5.9	96	0.00
15 T	Acrylonitrile	0.147	0.136	7.5	83	0.00
16 T	Acetone	0.183	0.158	13.7	79	0.03
17 T	Carbon Disulfide	1.630	1.689	-3.6	93	0.00
18 T	Methyl Acetate	0.410	0.361	12.0	79	0.00
19 T	Methyl tert-butyl Ether	0.826	0.822	0.5	88	0.00
20 T	Methylene Chloride	0.630	0.588	6.7	91	0.00
21 T	trans-1,2-Dichloroethene	0.561	0.583	-3.9	96	0.00
22 T	Diisopropyl ether	2.088	2.094	-0.3	91	0.00
23 T	Vinyl Acetate	1.312	1.318	-0.5	89	0.00
24 P	1,1-Dichloroethane	1.118	1.152	-3.0	94	0.00
25 T	2-Butanone	0.239	0.209	12.6	77	0.00
26 T	2,2-Dichloropropane	0.688	0.667	3.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.618	0.615	0.5	92	0.00
28 T	Bromochloromethane	0.496	0.527	-6.3	107	0.00
29 T	Tetrahydrofuran	0.137	0.124	9.5	81	0.00
30 C	Chloroform	1.032	1.044	-1.2#	94	0.00
31 T	Cyclohexane	1.073	1.093	-1.9	96	0.00
32 T	1,1,1-Trichloroethane	0.786	0.853	-8.5	98	0.00
33 S	1,2-Dichloroethane-d4	0.574	0.564	1.7	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.281	0.294	-4.6	94	0.00
36 T	1,1-Dichloropropene	0.520	0.576	-10.8	97	0.00
37 T	Ethyl Acetate	0.299	0.289	3.3	83	0.00
38 T	Carbon Tetrachloride	0.445	0.498	-11.9	97	0.00
39 T	Methylcyclohexane	0.572	0.653	-14.2	99	0.00
40 TM	Benzene	1.456	1.531	-5.2	94	0.00
41 T	Methacrylonitrile	0.179	0.183	-2.2	91	0.00
42 TM	1,2-Dichloroethane	0.456	0.471	-3.3	91	0.00
43 T	Isopropyl Acetate	0.550	0.550	0.0	86	0.00
44 TM	Trichloroethene	0.350	0.376	-7.4	95	0.00
45 C	1,2-Dichloropropane	0.390	0.407	-4.4#	92	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080619\
 Data File : VW011708.D
 Acq On : 05 Aug 2019 14:27
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 06 08:32:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.178	0.184	-3.4	91	0.00
47 T	Bromodichloromethane	0.467	0.494	-5.8	93	0.00
48 T	Methyl methacrylate	0.264	0.264	0.0	84	0.00
49 T	1,4-Dioxane	0.003	0.002	33.3#	78	0.04
50 S	Toluene-d8	1.148	1.216	-5.9	96	0.00
51 T	4-Methyl-2-Pentanone	0.286	0.280	2.1	83	0.00
52 CM	Toluene	0.879	0.942	-7.2#	94	0.00
53 T	t-1,3-Dichloropropene	0.487	0.507	-4.1	89	0.00
54 T	cis-1,3-Dichloropropene	0.579	0.608	-5.0	90	0.00
55 T	1,1,2-Trichloroethane	0.252	0.257	-2.0	89	0.00
56 T	Ethyl methacrylate	0.369	0.380	-3.0	86	0.00
57 T	1,3-Dichloropropane	0.477	0.486	-1.9	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.168	0.203	-20.8#	102	0.00
59 T	2-Hexanone	0.202	0.198	2.0	80	0.00
60 T	Dibromochloromethane	0.280	0.293	-4.6	90	0.00
61 T	1,2-Dibromoethane	0.239	0.239	0.0	87	0.00
62 S	4-Bromofluorobenzene	0.409	0.426	-4.2	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.326	0.356	-9.2	97	0.00
65 PM	Chlorobenzene	1.005	1.071	-6.6	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.350	0.374	-6.9	91	0.00
67 C	Ethyl Benzene	1.915	2.103	-9.8#	94	0.00
68 T	m/p-Xylenes	0.696	0.756	-8.6	93	0.00
69 T	o-Xylene	0.647	0.702	-8.5	92	0.00
70 T	Styrene	1.119	1.216	-8.7	92	0.00
71 P	Bromoform	0.179	0.190	-6.1	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.730	4.112	-10.2	95	0.00
74 T	N-amyl acetate	1.168	1.213	-3.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	0.702	0.723	-3.0	87	0.00
76 T	1,2,3-Trichloropropane	0.529	0.489	7.6	84	0.00
77 T	Bromobenzene	0.824	0.876	-6.3	91	0.00
78 T	n-propylbenzene	4.540	5.031	-10.8	95	0.00
79 T	2-Chlorotoluene	2.615	2.817	-7.7	94	0.00
80 T	1,3,5-Trimethylbenzene	3.127	3.458	-10.6	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.230	0.249	-8.3	88	0.00
82 T	4-Chlorotoluene	2.744	2.984	-8.7	94	0.00
83 T	tert-Butylbenzene	2.615	2.933	-12.2	95	0.00
84 T	1,2,4-Trimethylbenzene	3.150	3.440	-9.2	94	0.00
85 T	sec-Butylbenzene	3.731	4.170	-11.8	96	0.00
86 T	p-Isopropyltoluene	3.385	3.793	-12.1	96	0.00
87 T	1,3-Dichlorobenzene	1.617	1.720	-6.4	92	0.00
88 T	1,4-Dichlorobenzene	1.608	1.707	-6.2	93	0.00
89 T	n-Butylbenzene	3.378	3.818	-13.0	97	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080619\
Data File : VW011708.D
Acq On : 05 Aug 2019 14:27
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Aug 06 08:32:14 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:23:30 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.599	0.665	-11.0	94	0.00
91 T	1,2-Dichlorobenzene	1.430	1.508	-5.5	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.122	-1.7	85	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.081	-7.2	91	0.00
94 T	Hexachlorobutadiene	0.628	0.704	-12.1	97	0.00
95 T	Naphthalene	1.787	1.924	-7.7	85	0.00
96 T	1,2,3-Trichlorobenzene	0.883	0.942	-6.7	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6