

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080624\
 Data File : VW029836.D
 Acq On : 06 Aug 2024 12:35
 Operator : SY/MD
 Sample : VW0806SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Aug 07 02:27:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Aug 07 02:26:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	234136	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	226505	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	98221	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	48890	14.708	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.840%
7) Chloroethane-d5	2.899	69	42821	18.044	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.160%
11) 1,1-Dichloroethene-d2	4.021	65	25004	16.479	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.920%
21) 2-Butanone-d5	7.075	46	38248	41.983	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.960%
24) Chloroform-d	7.648	84	132561	21.871	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.480%
26) 1,2-Dichloroethane-d4	8.307	65	78211	22.655	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.640%
32) Benzene-d6	8.270	84	241223	19.710	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.840%
36) 1,2-Dichloropropane-d6	9.270	67	83116	21.269	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.080%
41) Toluene-d8	10.325	98	208893	19.196	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.800%
43) trans-1,3-Dichloroprop...	10.575	79	29345	18.558	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.240%
47) 2-Hexanone-d5	10.922	63	23917	37.117	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.240%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	64820	22.339	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.360%
66) 1,2-Dichlorobenzene-d4	13.848	152	70259	23.037	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW080624\
Data File : VW029836.D
Acq On : 06 Aug 2024 12:35
Operator : SY/MD
Sample : VW0806SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Aug 07 02:27:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 07 02:26:16 2024
Response via : Initial Calibration

