

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080720\
 Data File : VW016069.D
 Acq On : 07 Aug 2020 11:48
 Operator : SY/VA
 Sample : L3579-04
 Misc : 5.90G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFWP7

Quant Time: Aug 08 04:12:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 08 03:18:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	362851	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	337296	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	150226	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	68569	13.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.56%
7) Chloroethane-d5	2.89	69	59841	14.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.56%
10) 1,1-Dichloroethene-d2	4.01	63	117610	11.36	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.44%
20) 2-Butanone-d5	7.09	46	41143	30.59	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.18%
24) Chloroform-d	7.65	84	182141	17.87	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	71.48%
26) 1,2-Dichloroethane-d4	8.31	65	93573	16.73	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	66.92%#
29) Benzene-d6	8.27	84	358201	17.45	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	69.80%
33) 1,2-Dichloropropane-d6	9.27	67	107749	17.51	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	70.04%
37) Toluene-d8	10.32	98	326252	17.54	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	70.16%
38) trans-1,3-Dichloropropene-	10.58	79	42491	16.20	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	64.80%
39) 2-Hexanone-d5	10.93	63	33781	32.96	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	65.92%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	85552	18.08	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	72.32%
61) 1,2-Dichlorobenzene-d4	13.85	152	119798	20.45	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	81.80%

Target Compounds					Ovalue
14) Carbon disulfide	4.39	76	3949	0.268 ug/L	100
47) Tetrachloroethene	10.87	164	37039	8.805 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW080720\
Data File : VW016069.D
Acq On : 07 Aug 2020 11:48
Operator : SY/VA
Sample : L3579-04
Misc : 5.90G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFWP7

Quant Time: Aug 08 04:12:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 08 03:18:33 2020
Response via : Initial Calibration

