

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW080818\
 Data File : VW004513.D
 Acq On : 08 Aug 2018 00:38
 Operator : SY/AP
 Sample : J4044-02
 Misc : 10.59G/5ML/MSVOA W/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JTY-SB-04-1-2

Quant Time: Aug 08 09:06:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 08:22:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	531851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	919089	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	866089	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	381746	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	117685	19.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	39.98%	
35) Dibromofluoromethane	7.88	113	109039	17.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	35.58%	
50) Toluene-d8	10.33	98	313159	13.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	26.80%	
62) 4-Bromofluorobenzene	12.63	95	105540	12.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	25.08%	
Target Compounds						
16) Acetone	4.14	43	51034	34.71	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW080818\
Data File : VW004513.D
Acq On : 08 Aug 2018 00:38
Operator : SY/AP
Sample : J4044-02
Misc : 10.59G/5ML/MSVOA W/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JTY-SB-04-1-2

Quant Time: Aug 08 09:06:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 08:22:01 2018
Response via : Initial Calibration

