

Data File : VW011769.D
Acq On : 08 Aug 2019 12:31
Operator : SY/VA
Sample : K4215-08
Misc : 5.67G/10ML/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ETOB9

Quant Time: Aug 09 02:30:31 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM073019S.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 09 02:27:11 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	268480	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	165534	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	212190	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.50	65	206	0.05	ug/L	0.14
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.20%#
7) Chloroethane-d5	2.89	69	65177	23.94	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.76%
10) 1,1-Dichloroethene-d2	4.04	63	498541	57.52	ug/L	0.02
Spiked Amount	25.000	Range	45 - 110	Recovery	=	230.08%#
20) 2-Butanone-d5	7.06	46	250	0.15	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.30%#
24) Chloroform-d	7.66	84	161918	21.96	ug/L	0.01
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.84%
26) 1,2-Dichloroethane-d4	8.31	65	102544	23.50	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.00%
29) Benzene-d6	8.28	84	346384	33.28	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	133.12%
33) 1,2-Dichloropropane-d6	9.29	67	220857	65.73	ug/L	0.01
Spiked Amount	25.000	Range	70 - 120	Recovery	=	262.92%#
37) Toluene-d8	10.33	98	647378	70.07	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	280.28%#
38) trans-1,3-Dichloropropene-	10.58	79	35670	24.22	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.88%
39) 2-Hexanone-d5	10.93	63	58477	74.50	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	149.00%#
48) 1,1,2,2-Tetrachloroethane-	12.68	84	930622	333.48	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 120	Recovery	=	1333.92%#
61) 1,2-Dichlorobenzene-d4	13.87	152	114130	13.96	ug/L	0.02
Spiked Amount	25.000	Range	75 - 120	Recovery	=	55.84%#

Target Compounds

						Qvalue
5) Vinyl chloride	2.35	62	19552091	3915.480	ug/L	93
11) 1,1,2-Trichloro-1,2,2-trif	4.04	101	2952	0.765	ug/L #	23
12) 1,1-Dichloroethene	4.04	96	561207	145.539	ug/L #	82
13) Acetone	3.84	43	17795	9.761	ug/L	49
14) Carbon disulfide	4.40	76	70297	5.597	ug/L	100
16) Methylene chloride	4.92	84	2932533	699.929	ug/L	98
18) trans-1,2-Dichloroethene	5.43	96	1545959	388.046	ug/L	99
19) 1,1-Dichloroethane	6.23	63	15605	1.961	ug/L #	44
21) 2-Butanone	6.89	43	71734	31.382	ug/L	96
25) Chloroform	7.65	83	71505	9.748	ug/L #	20
30) Cyclohexane	7.96	56	353143	61.373	ug/L	85
34) Benzene	8.33	78	1459406	125.809	ug/L	100
35) Trichloroethene	9.10	95	4254316	1444.638	ug/L	99
36) Methylcyclohexane	9.34	83	628550	117.652	ug/L	91
40) 1,2-Dichloropropane	9.34	63	8444	2.727	ug/L #	1
41) Bromodichloromethane	9.63	83	50520	13.826	ug/L #	45
43) 4-Methyl-2-pentanone	10.21	43	344329	122.733	ug/L	98

Data File : VW011769.D
Acq On : 08 Aug 2019 12:31
Operator : SY/VA
Sample : K4215-08
Misc : 5.67G/10ML/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ETOB9

Quant Time: Aug 09 02:30:31 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM073019S.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 09 02:27:11 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) Toluene	10.39	91	19496393	1654.284	ug/L #	69
46) 1,1,2-Trichloroethane	10.77	97	1090859	516.724	ug/L #	24
47) Tetrachloroethene	10.87	164	351176	156.121	ug/L	94
49) 2-Hexanone	10.95	43	7338079	3394.945	ug/L #	29
51) 1,2-Dibromoethane	11.24	107	3083	1.501	ug/L #	1
52) Chlorobenzene	11.66	112	27197	3.825	ug/L #	5
53) Ethylbenzene	11.74	91	13693863	1041.918	ug/L	93
54) m,p-Xylene	11.85	106	14294200	2990.967	ug/L #	44
55) o-xylene	12.18	106	5699580	1267.219	ug/L	86
56) Styrene	12.18	104	288268	37.699	ug/L #	1
57) Isopropylbenzene	12.49	105	2791371	225.591	ug/L #	92
58) 1,1,2,2-Tetrachloroethane	12.71	83	1491586	531.677	ug/L #	74
59) 1,2,3-Trichloropropane	12.81	75	79598	36.787	ug/L #	1
66) 1,2-Dibromo-3-chloropropan	14.50	75	4192	3.106	ug/L #	18
69) Naphthalene	15.36	128	99480	5.787	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW011769.D
Acq On : 08 Aug 2019 12:31
Operator : SY/VA
Sample : K4215-08
Misc : 5.67G/10ML/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ETOB9

Quant Time: Aug 09 02:30:31 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM073019S.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 09 02:27:11 2019
Response via : Initial Calibration





#5

Vinyl chloride

Concen: 3915.480 ug/L

RT: 2.35 min Scan# 74

Delta R.T. -0.01 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31

Instrument :

MSVOA_W

ClientSampled :





ETOB9

Tgt Ion: 62 Resp:19552091

Ion Ratio Lower Upper

62 100

64 37.5 23.6 43.8



#11

1,1,2-Trichloro-1,2,2-trifluoroethane

Concen: 0.765 ug/L

RT: 4.04 min Scan# 351

Delta R.T. -0.02 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31




Tgt Ion:101 Resp: 2952


Ion Ratio Lower Upper

101 100

85 7.5 37.0 55.4#

151 0.0 60.8 91.2#









#12
 1,1-Dichloroethene
 Concen: 145.539 ug/L
 RT: 4.04 min Scan# 351
 Delta R.T. 0.01 min
 Lab File: VW011769.D
 Acq: 08 Aug 2019 12:31




Instrument :
 MSVOA_W
 ClientSampled :
 ETOB9


Tgt Ion: 96 Resp: 561207
 Ion Ratio Lower Upper
 96 100
 61 195.9 135.0 250.8
 63 78.6 89.2 165.8#



#13
 Acetone
 Concen: 9.761 ug/L
 RT: 3.84 min Scan# 318
 Delta R.T. -0.27 min
 Lab File: VW011769.D
 Acq: 08 Aug 2019 12:31

Tgt Ion: 43 Resp: 17795
 Ion Ratio Lower Upper
 43 100
 58 0.2 0.0 53.4





#14

Carbon disulfide

Concen: 5.597 ug/L

RT: 4.40 min Scan# 409

Delta R.T. 0.01 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31

Instrument :

MSVOA_W

ClientSampled :





ETOB9

Tgt Ion: 76 Resp: 70297

Ion Ratio Lower Upper

76 100

78 9.0 7.0 10.6



#16

Methylene chloride

Concen: 699.929 ug/L

RT: 4.92 min Scan# 495

Delta R.T. 0.01 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31




Tgt Ion: 84 Resp: 2932533


Ion Ratio Lower Upper

84 100

49 152.5 106.8 198.4

86 65.0 42.8 79.4









#18
trans-1,2-Dichloroethene
Concen: 388.046 ug/L
RT: 5.43 min Scan# 579
Delta R.T. 0.01 min
Lab File: VW011769.D
Acq: 08 Aug 2019 12:31




Instrument :
MSVOA_W
ClientSampleId :
ETOB9


Tgt Ion: 96 Resp: 1545959
Ion Ratio Lower Upper
96 100
61 164.9 114.2 212.0
98 63.7 45.6 84.6



#19
1,1-Dichloroethane
Concen: 1.961 ug/L
RT: 6.23 min Scan# 710
Delta R.T. 0.02 min
Lab File: VW011769.D
Acq: 08 Aug 2019 12:31

Tgt Ion: 63 Resp: 15605
Ion Ratio Lower Upper
63 100
65 70.4 21.7 40.3#
83 18.4 8.4 15.6#





#21

2-Butanone

Concen: 31.382 ug/L

RT: 6.89 min Scan# 818

Delta R.T. -0.27 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31

Instrument :

MSVOA_W

ClientSampled :





ETOB9

Tgt Ion: 43 Resp: 71734

Ion Ratio Lower Upper

43 100

72 23.1 10.7 32.1



#25

Chloroform

Concen: 9.748 ug/L

RT: 7.65 min Scan# 942

Delta R.T. -0.03 min

Lab File: VW011769.D




Acq: 08 Aug 2019 12:31


Tgt Ion: 83 Resp: 71505

Ion Ratio Lower Upper

83 100

85 2.1 45.4 84.2#









#30
Cyclohexane
Concen: 61.373 ug/L
RT: 7.96 min Scan# 993
Delta R.T. 0.00 min
Lab File: VW011769.D
Acq: 08 Aug 2019 12:31




Instrument :
MSVOA_W
ClientSampled :
ETOB9


Tgt Ion: 56 Resp: 353143
Ion Ratio Lower Upper
56 100
69 32.6 22.4 33.6
84 61.3 60.6 91.0



#34
Benzene
Concen: 125.809 ug/L
RT: 8.33 min Scan# 1054
Delta R.T. 0.01 min
Lab File: VW011769.D
Acq: 08 Aug 2019 12:31

Tgt Ion: 78 Resp: 1459406





#35

Trichloroethene

Concen: 1444.638 ug/L

RT: 9.10 min Scan# 1180

Delta R.T. 0.01 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31

Instrument :

MSVOA_W

ClientSampleId :

ETOB9

Tgt Ion: 95 Resp: 4254316



Ion Ratio Lower Upper

95 100

97 65.2 44.8 83.2

132 92.9 65.0 120.6

130 96.3 68.0 126.4





Abundance

Ion 95.00 (94.70 to 95.70): VW

Ion 97.00 (96.70 to 97.70): VW

Ion 132.00 (131.70 to 132.70): VW

Ion 130.00 (129.70 to 130.70): VW



#36

Methylcyclohexane

Concen: 117.652 ug/L

RT: 9.34 min Scan# 1220

Delta R.T. 0.01 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31



Tgt Ion: 83 Resp: 628550

Ion Ratio Lower Upper

83 100

55 100.6 72.6 108.8

98 43.9 37.8 56.6





Abundance

Ion 83.00 (82.70 to 83.70): VW

Ion 55.00 (54.70 to 55.70): VW

Ion 98.00 (97.70 to 98.70): VW





#40

1,2-Dichloropropane

Concen: 2.727 ug/L

RT: 9.34 min Scan# 1220

Delta R.T. -0.02 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31

Instrument :

MSVOA_W

ClientSampled :





ETOB9

Tgt Ion: 63 Resp: 8444

Ion Ratio Lower Upper

63 100

112 110.9 3.1 4.7#



#41

Bromodichloromethane

Concen: 13.826 ug/L

RT: 9.63 min Scan# 1267

Delta R.T. -0.02 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31




Tgt Ion: 83 Resp: 50520


Ion Ratio Lower Upper

83 100

85 17.4 44.0 81.6#

127 0.0 6.5 9.7#









#43
4-Methyl-2-pentanone
Concen: 122.733 ug/L
RT: 10.21 min Scan# 1363
Delta R.T. 0.01 min
Lab File: VW011769.D
Acq: 08 Aug 2019 12:31




Instrument :
MSVOA_W
ClientSampled :
ETOB9


Tgt Ion: 43 Resp: 344329
Ion Ratio Lower Upper
43 100
58 33.6 28.0 42.0
100 11.5 9.4 14.2



#44
Toluene
Concen: 1654.284 ug/L
RT: 10.39 min Scan# 1392
Delta R.T. 0.00 min
Lab File: VW011769.D
Acq: 08 Aug 2019 12:31

Tgt Ion: 91 Resp: 19496393
Ion Ratio Lower Upper
91 100
92 81.4 40.9 76.0#









#46
 1,1,2-Trichloroethane
 Concen: 516.724 ug/L
 RT: 10.77 min Scan# 1455
 Delta R.T. -0.01 min
 Lab File: VW011769.D
 Acq: 08 Aug 2019 12:31

Instrument :
 MSVOA_W
 ClientSampleId :
 ETOB9

Tgt Ion: 97 Resp: 1090859
 Ion Ratio Lower Upper
 97 100
 83 9.6 63.8 118.6#
 85 17.1 39.3 72.9#
 99 0.3 43.5 80.7#




Abundance Ion 97.00 (96.70 to 97.70): VW
 Ion 83.00 (82.70 to 83.70): VW
 Ion 85.00 (84.70 to 85.70): VW
 Ion 99.00 (98.70 to 99.70): VW






#47
 Tetrachloroethene
 Concen: 156.121 ug/L
 RT: 10.87 min Scan# 1470
 Delta R.T. 0.01 min
 Lab File: VW011769.D
 Acq: 08 Aug 2019 12:31

Tgt Ion: 164 Resp: 351176
 Ion Ratio Lower Upper
 164 100
 129 97.8 64.0 119.0
 131 93.5 58.6 108.8
 166 128.3 91.7 170.3



Abundance Ion 164.00 (163.70 to 164.70): V
 Ion 129.00 (128.70 to 129.70): V
 Ion 131.00 (130.70 to 131.70): V
 Ion 166.00 (165.70 to 166.70): V









#49
2-Hexanone
Concen: 3394.945 ug/L
RT: 10.95 min Scan# 1484
Delta R.T. -0.02 min
Lab File: VW011769.D
Acq: 08 Aug 2019 12:31




Instrument :
MSVOA_W
ClientSampleId :
ETOB9


Tgt Ion: 43 Resp: 7338079
Ion Ratio Lower Upper
43 100
58 2.8 38.6 57.8#
57 63.4 13.0 19.6#
100 0.0 7.5 11.3#



#51
1,2-Dibromoethane
Concen: 1.501 ug/L
RT: 11.24 min Scan# 1532
Delta R.T. 0.01 min
Lab File: VW011769.D
Acq: 08 Aug 2019 12:31

Tgt Ion: 107 Resp: 3083
Ion Ratio Lower Upper
107 100
109 835.2 66.1 122.8#
188 0.0 2.9 4.3#





#52

Chlorobenzene

Concen: 3.825 ug/L

RT: 11.66 min Scan# 1601

Delta R.T. 0.01 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31

Instrument :

MSVOA_W

ClientSampleId :

ETOB9





Tgt Ion: 112 Resp: 27197

Ion Ratio Lower Upper

112 100

77 170.1 51.0 76.4#

114 25.7 22.4 41.6



#53

Ethylbenzene

Concen: 1041.918 ug/L

RT: 11.74 min Scan# 1613

Delta R.T. 0.01 min

Lab File: VW011769.D




Acq: 08 Aug 2019 12:31


Tgt Ion: 91 Resp: 13693863

Ion Ratio Lower Upper

91 100

106 33.2 20.6 38.4





#54

m,p-Xylene

Concen: 2990.967 ug/L

RT: 11.85 min Scan# 1631

Delta R.T. 0.01 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31

Instrument :

MSVOA_W

ClientSampleId :


ETOB9

Tgt Ion:106 Resp:14294200

Ion Ratio Lower Upper

106 100

91 125.6 149.6 277.8#



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VW

8000000

6000000



4000000

2000000

0

11.80 11.85 11.90

Time-->



#55

o-xylene

Concen: 1267.219 ug/L

RT: 12.18 min Scan# 1686

Delta R.T. 0.02 min

Lab File: VW011769.D


Acq: 08 Aug 2019 12:31

Tgt Ion:106 Resp: 5699580

Ion Ratio Lower Upper

106 100

91 202.7 158.3 293.9



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VW

6000000


4000000


2000000

0

12.10 12.18 12.30

Time-->





#56

Styrene

Concen: 37.699 ug/L

RT: 12.18 min Scan# 1686

Delta R.T. 0.00 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31

Instrument :

MSVOA_W

ClientSampled :





ETOB9

Tgt Ion:104 Resp: 288268

Ion Ratio Lower Upper

104 100

78 327.2 32.9 61.1#



#57

Isopropylbenzene

Concen: 225.591 ug/L

RT: 12.49 min Scan# 1736

Delta R.T. 0.02 min

Lab File: VW011769.D

Acq: 08 Aug 2019 12:31




Tgt Ion:105 Resp: 2791371


Ion Ratio Lower Upper

105 100

120 25.6 20.6 30.8

77 25.7 13.4 20.2#









#58
 1,1,2,2-Tetrachloroethane
 Concen: 531.677 ug/L
 RT: 12.71 min Scan# 1773
 Delta R.T. 0.00 min
 Lab File: VW011769.D
 Acq: 08 Aug 2019 12:31




Instrument :
 MSVOA_W
 ClientSampled :
 ETOB9


Tgt Ion: 83 Resp: 1491586
 Ion Ratio Lower Upper
 83 100
 85 41.9 43.5 80.9#
 131 0.0 7.5 11.3#



#59
 1,2,3-Trichloropropane
 Concen: 36.787 ug/L
 RT: 12.81 min Scan# 1789
 Delta R.T. 0.04 min
 Lab File: VW011769.D
 Acq: 08 Aug 2019 12:31

Tgt Ion: 75 Resp: 79598
 Ion Ratio Lower Upper
 75 100
 77 1078.1 25.1 37.7#









#66
 1,2-Dibromo-3-chloropropane
 Concen: 3.106 ug/L
 RT: 14.50 min Scan# 2066
 Delta R.T. 0.01 min
 Lab File: VW011769.D
 Acq: 08 Aug 2019 12:31

Instrument :
 MSVOA_W
 ClientSampled :
 ETOB9

Tgt Ion: 75 Resp: 4192
 Ion Ratio Lower Upper
 75 100
 157 0.0 79.3 118.9#
 155 24.9 59.9 89.9#



#69
 Naphthalene
 Concen: 5.787 ug/L
 RT: 15.36 min Scan# 2208
 Delta R.T. 0.00 min
 Lab File: VW011769.D
 Acq: 08 Aug 2019 12:31

Tgt Ion: 128 Resp: 99480
 Ion Ratio Lower Upper
 128 100
 129 10.9 8.7 13.1
 127 13.8 10.6 16.0

