

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081018\
 Data File : VW004540.D
 Acq On : 08 Aug 2018 21:09
 Operator : SY/AP
 Sample : J4310-07RE
 Misc : 3.81G/5ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C-4RE

Quant Time: Aug 09 05:41:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 08:22:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	411375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	644708	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	453543	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	95762	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	298177	65.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	130.96%	
35) Dibromofluoromethane	7.88	113	270050	62.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	125.64%	
50) Toluene-d8	10.33	98	705896	43.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.14%	
62) 4-Bromofluorobenzene	12.62	95	121660	20.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	41.22%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081018\
Data File : VW004540.D
Acq On : 08 Aug 2018 21:09
Operator : SY/AP
Sample : J4310-07RE
Misc : 3.81G/5ML/MSVOA W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C-4RE

Quant Time: Aug 09 05:41:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 08:22:01 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.95 min Scan# 1044

Delta R.T. 0.00 min

Lab File: VW004540.D

Acq: 08 Aug 2018 21:09

Instrument :

MSVOA_W

ClientSampled :


C-4RE

Tot Ion:168 Resp: 411375

Ion Ratio Lower Upper

168 100

99 55.1 44.1 66.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95

200000



150000

100000

50000

0

Time--> 7.80 7.90 8.00 8.10



#33

1,2-Dichloroethane-d4

Concen: 65.48 ug/l

RT: 8.31 min Scan# 1102

Delta R.T. -0.01 min

Lab File: VW004540.D


Acq: 08 Aug 2018 21:09

Tgt Ion: 65 Resp: 298177

Ion Ratio Lower Upper

65 100

67 53.7 0.0 109.0



Abundance Ion 64.90 (64.60 to 65.60): VW

Ion 66.90 (66.60 to 67.60): VW

8.31


150000


100000

50000

0

Time--> 8.20 8.30 8.40





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.85 min Scan# 1191

Delta R.T. 0.00 min

Lab File: VW004540.D

Acq: 08 Aug 2018 21:09

Instrument :

MSVOA_W

ClientSampleId :

C-4RE


Tot Ion:114 Resp: 644708

Ion Ratio Lower Upper

114 100

63 20.8 0.0 42.4

88 15.8 0.0 31.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VW

Ion 87.90 (87.60 to 88.60): VW

400000

300000

200000

100000



0

8.80

8.85

8.90

Time-->



#35

Dibromofluoromethane

Concen: 62.82 ug/l

RT: 7.88 min Scan# 1033

Delta R.T. 0.00 min

Lab File: VW004540.D

Acq: 08 Aug 2018 21:09


Tgt Ion:113 Resp: 270050

Ion Ratio Lower Upper

113 100

111 102.3 82.9 124.3

192 20.1 15.8 23.6



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V

150000

100000

50000


0


7.80

7.88

7.90

Time-->





#50

Toluene-d8

Concen: 43.07 ug/l

RT: 10.33 min Scan# 1434

Delta R.T. 0.00 min

Lab File: VW004540.D

Acq: 08 Aug 2018 21:09

Instrument :

MSVOA_W

ClientSampleId :





C-4RE

Tot Ion: 98 Resp: 705896

Ion Ratio Lower Upper

98 100

100 63.2 51.4 77.2



#62

4-Bromofluorobenzene

Concen: 20.61 ug/l

RT: 12.62 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VW004540.D

Acq: 08 Aug 2018 21:09




Tgt Ion: 95 Resp: 121660


Ion Ratio Lower Upper

95 100

174 82.9 0.0 160.4

176 79.7 0.0 153.2









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.63 min Scan# 1648
Delta R.T. 0.00 min
Lab File: VW004540.D
Acq: 08 Aug 2018 21:09

Instrument :
MSVOA_W
ClientSampleId :
C-4RE

Tot Ion:117 Resp: 453543
Ion Ratio Lower Upper
117 100
82 57.4 45.3 67.9
119 31.6 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.57 min Scan# 1965
Delta R.T. 0.00 min
Lab File: VW004540.D
Acq: 08 Aug 2018 21:09

Tgt Ion:152 Resp: 95762
Ion Ratio Lower Upper
152 100
115 58.1 29.1 87.3
150 156.1 0.0 349.6

