

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081023\
 Data File : VW026690.D
 Acq On : 10 Aug 2023 20:32
 Operator : SY/MD
 Sample : 03967-02MS
 Misc : 5.80g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A48K2MS

Quant Time: Aug 11 02:04:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 11 02:00:25 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	651262	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	525497	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	163590	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	272998	25.499	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 102.000%	
7) Chloroethane-d5	2.893	69	212185	26.653	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 106.600%	
11) 1,1-Dichloroethene-d2	4.021	65	131308	24.937	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	= 99.760%	
21) 2-Butanone-d5	7.075	46	122571	39.833	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 79.660%	
24) Chloroform-d	7.648	84	589435	26.966	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 107.880%	
26) 1,2-Dichloroethane-d4	8.307	65	336525	27.898	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 111.600%	
32) Benzene-d6	8.270	84	1185777	32.741	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 130.960%	
36) 1,2-Dichloropropane-d6	9.276	67	371418	32.923	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 131.680%#	
41) Toluene-d8	10.318	98	960614	30.899	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 123.600%	
43) trans-1,3-Dichloroprop...	10.575	79	109707	23.553	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 94.200%	
47) 2-Hexanone-d5	10.922	63	56584	47.734	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 95.460%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	248025	27.840	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 111.360%	
66) 1,2-Dichlorobenzene-d4	13.848	152	151521	25.878	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 103.520%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.015	85	195927	22.127	ug/L	96
3) Chloromethane	2.228	50	263088	21.729	ug/L	96
5) Vinyl chloride	2.375	62	265081	22.902	ug/L	97
6) Bromomethane	2.783	94	103706	15.378	ug/L	99
8) Chloroethane	2.930	64	155734	22.455	ug/L	99
9) Trichlorofluoromethane	3.259	101	214653	22.597	ug/L	100
10) 1,1,2-Trichloro-1,2,2-...	4.070	101	207090	20.183	ug/L	99
12) 1,1-Dichloroethene	4.045	96	212170	22.041	ug/L	88
13) Acetone	4.118	43	99658	32.573	ug/L	97
14) Carbon disulfide	4.393	76	572383	16.528	ug/L	99
15) Methyl Acetate	4.673	43	100324	16.370	ug/L	99
16) Methylene chloride	4.917	84	241692	19.185	ug/L	95
17) trans-1,2-Dichloroethene	5.429	96	233908	21.775	ug/L	99
18) Methyl tert-butyl Ether	5.423	73	416455	25.623	ug/L	99
19) 1,1-Dichloroethane	6.216	63	512860	23.198	ug/L	98
20) cis-1,2-Dichloroethene	7.173	96	258093	22.354	ug/L	97
22) 2-Butanone	7.173	43	204065	45.258	ug/L	95
23) Bromochloromethane	7.514	128	116393	23.337	ug/L	97
25) Chloroform	7.673	83	476246	22.298	ug/L	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081023\
 Data File : VW026690.D
 Acq On : 10 Aug 2023 20:32
 Operator : SY/MD
 Sample : 03967-02MS
 Misc : 5.80g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A48K2MS


Quant Time: Aug 11 02:04:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 11 02:00:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
27) 1,2-Dichloroethane	8.398	62	347349	23.312	ug/L	99
29) Cyclohexane	7.953	56	330807	20.356	ug/L	99
30) 1,1,1-Trichloroethane	7.874	97	362387	26.423	ug/L	99
31) Carbon tetrachloride	8.069	117	306358	24.431	ug/L	95
33) Benzene	8.325	78	1042965	27.254	ug/L	100
34) Trichloroethene	9.093	95	240672	24.693	ug/L	96
35) Methylcyclohexane	9.331	83	229916	13.789	ug/L	100
37) 1,2-Dichloropropane	9.367	63	284741	27.165	ug/L	99
38) Bromodichloromethane	9.642	83	318466	25.899	ug/L	99
39) cis-1,3-Dichloropropene	10.075	75	289654	19.209	ug/L	98
40) 4-Methyl-2-pentanone	10.209	43	399860	58.646	ug/L	99
42) Toluene	10.386	91	1000724	25.851	ug/L	99
44) trans-1,3-Dichloropropene	10.605	75	256829	19.788	ug/L	95
45) 1,1,2-Trichloroethane	10.788	97	193867	27.043	ug/L	98
46) Tetrachloroethene	10.855	164	141300	19.955	ug/L	98
48) 2-Hexanone	10.965	43	253467	54.633	ug/L	95
49) Dibromochloromethane	11.123	129	183253	24.441	ug/L	97
50) 1,2-Dibromoethane	11.233	107	159105	23.654	ug/L	99
51) Chlorobenzene	11.654	112	524462	21.448	ug/L	99
52) Ethylbenzene	11.727	91	938355	21.615	ug/L	99
53) m,p-Xylene	11.836	106	339452	21.107	ug/L	93
54) o-Xylene	12.166	106	317210	21.134	ug/L	99
55) Styrene	12.178	104	458971	17.586	ug/L	98
57) 1,1,2,2-Tetrachloroethane	12.708	83	208249	22.938	ug/L	98
59) Bromoform	12.349	173	91745	35.801	ug/L	97
60) Isopropylbenzene	12.458	105	723205	29.735	ug/L	99
61) 1,2,3-Trichloropropane	12.763	75	169927	42.822	ug/L	98
62) 1,3,5-Trimethylbenzene	12.940	105	493461	25.669	ug/L	99
63) 1,2,4-Trimethylbenzene	13.245	105	475538	25.243	ug/L	99
64) 1,3-Dichlorobenzene	13.495	146	235745	22.287	ug/L	94
65) 1,4-Dichlorobenzene	13.574	146	242440	21.963	ug/L	97
67) 1,2-Dichlorobenzene	13.867	146	207661	21.589	ug/L	97
68) 1,2-Dibromo-3-chloropr...	14.476	75	30188	31.556	ug/L	90
69) 1,3,5-Trichlorobenzene	14.623	180	73800	10.327	ug/L	97
70) 1,2,4-trichlorobenzene	15.129	180	51503	8.500	ug/L	94
71) Naphthalene	15.360	128	131715	11.396	ug/L	99
72) 1,2,3-Trichlorobenzene	15.549	180	41510	7.596	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 11 02:04:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 11 02:00:25 2023
Response via : Initial Calibration









#2
Dichlorodifluoromethane
Concen: 22.127 ug/L
RT: 2.015 min Scan# 21
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 85 Resp: 195927
Ion Ratio Lower Upper
85 100
87 26.5 19.5 29.3



#3
Chloromethane
Concen: 21.729 ug/L
RT: 2.228 min Scan# 56
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 50 Resp: 263088
Ion Ratio Lower Upper
50 100
52 30.8 23.0 42.8









#5
 Vinyl chloride
 Concen: 22.902 ug/L
 RT: 2.375 min Scan# 80
 Delta R.T. -0.000 min
 Lab File: VW026690.D
 Acq: 10 Aug 2023 20:32




Instrument :
 MSVOA_W
 ClientSampleId :
 A48K2MS


Tgt Ion: 62 Resp: 265081
 Ion Ratio Lower Upper
 62 100
 64 34.2 22.7 42.3



#6
 Bromomethane
 Concen: 15.378 ug/L
 RT: 2.783 min Scan# 147
 Delta R.T. -0.000 min
 Lab File: VW026690.D
 Acq: 10 Aug 2023 20:32

Tgt Ion: 94 Resp: 103706
 Ion Ratio Lower Upper
 94 100
 96 93.4 64.8 120.3









#8
Chloroethane
Concen: 22.455 ug/L
RT: 2.930 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 64 Resp: 155734
Ion Ratio Lower Upper
64 100
66 31.8 21.7 40.3



#9
Trichlorofluoromethane
Concen: 22.597 ug/L
RT: 3.259 min Scan# 225
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 101 Resp: 214653
Ion Ratio Lower Upper
101 100
103 64.6 51.9 77.9









#10
 1,1,2-Trichloro-1,2,2-trifluoroethane
 Concen: 20.183 ug/L
 RT: 4.070 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: VW026690.D
 Acq: 10 Aug 2023 20:32




Instrument :
 MSVOA_W
 ClientSampleId :
 A48K2MS


Tgt Ion:101 Resp: 207090
 Ion Ratio Lower Upper
 101 100
 85 40.5 32.8 49.2
 151 67.1 54.4 81.6



#12
 1,1-Dichloroethene
 Concen: 22.041 ug/L
 RT: 4.045 min Scan# 354
 Delta R.T. -0.006 min
 Lab File: VW026690.D
 Acq: 10 Aug 2023 20:32

Tgt Ion: 96 Resp: 212170
 Ion Ratio Lower Upper
 96 100
 61 195.1 133.2 247.4
 63 184.3 109.0 202.4









#13
Acetone
Concen: 32.573 ug/L
RT: 4.118 min Scan# 366
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 43 Resp: 99658
Ion Ratio Lower Upper
43 100
58 30.1 0.0 57.4



#14
Carbon disulfide
Concen: 16.528 ug/L
RT: 4.393 min Scan# 411
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 76 Resp: 572383
Ion Ratio Lower Upper
76 100
78 9.2 7.1 10.7









#15
Methyl Acetate
Concen: 16.370 ug/L
RT: 4.673 min Scan# 41
Delta R.T. 0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 43 Resp: 100324
Ion Ratio Lower Upper
43 100
74 20.2 16.4 24.6



#16
Methylene chloride
Concen: 19.185 ug/L
RT: 4.917 min Scan# 497
Delta R.T. -0.012 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 84 Resp: 241692
Ion Ratio Lower Upper
84 100
86 63.7 43.5 80.9
49 145.3 96.0 178.2









#17
trans-1,2-Dichloroethene
Concen: 21.775 ug/L
RT: 5.429 min Scan# 580
Delta R.T. 0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 96 Resp: 233908
Ion Ratio Lower Upper
96 100
61 154.3 109.4 203.2
98 63.4 43.6 81.0



#18
Methyl tert-butyl Ether
Concen: 25.623 ug/L
RT: 5.423 min Scan# 580
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 73 Resp: 416455
Ion Ratio Lower Upper
73 100
43 22.5 18.0 27.0
57 23.7 19.3 28.9





#19

1,1-Dichloroethane

Concen: 23.198 ug/L

RT: 6.216 min Scan# 710

Delta R.T. -0.000 min

Lab File: VW026690.D

Acq: 10 Aug 2023 20:32

Instrument :

MSVOA_W

ClientSampleId :

A48K2MS





Tgt Ion: 63 Resp: 512860

Ion Ratio Lower Upper

63 100

65 32.6 21.8 40.6

83 12.0 9.0 16.6



#20

cis-1,2-Dichloroethene

Concen: 22.354 ug/L

RT: 7.173 min Scan# 867

Delta R.T. 0.006 min

Lab File: VW026690.D

Acq: 10 Aug 2023 20:32




Tgt Ion: 96 Resp: 258093


Ion Ratio Lower Upper

96 100

61 140.3 101.2 188.0

98 60.3 43.8 81.3









#22
2-Butanone
Concen: 45.258 ug/L
RT: 7.173 min Scan# 867
Delta R.T. 0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 43 Resp: 204065
Ion Ratio Lower Upper
43 100
72 21.4 11.8 35.4



#23
Bromochloromethane
Concen: 23.337 ug/L
RT: 7.514 min Scan# 923
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 128 Resp: 116393
Ion Ratio Lower Upper
128 100
49 206.8 143.1 265.7
130 127.1 91.8 170.4
51 63.1 53.0 79.4









#25
Chloroform
Concen: 22.298 ug/L
RT: 7.673 min Scan# 949
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 83 Resp: 476246
Ion Ratio Lower Upper
83 100
85 65.7 45.4 84.2



#27
1,2-Dichloroethane
Concen: 23.312 ug/L
RT: 8.398 min Scan# 1068
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 62 Resp: 347349
Ion Ratio Lower Upper
62 100
98 8.6 7.1 10.7





#29

Cyclohexane

Concen: 20.356 ug/L

RT: 7.953 min Scan# 996

Delta R.T. -0.006 min

Lab File: VW026690.D

Acq: 10 Aug 2023 20:32

Instrument :

MSVOA_W

ClientSampleId :

A48K2MS





Tgt Ion: 56 Resp: 330807

Ion Ratio Lower Upper

56 100

69 29.9 23.2 34.8

84 79.1 62.7 94.1



#30

1,1,1-Trichloroethane

Concen: 26.423 ug/L

RT: 7.874 min Scan# 982

Delta R.T. -0.000 min

Lab File: VW026690.D

Acq: 10 Aug 2023 20:32




Tgt Ion: 97 Resp: 362387


Ion Ratio Lower Upper

97 100

99 64.3 51.2 76.8

61 49.6 39.2 58.8









#31
Carbon tetrachloride
Concen: 24.431 ug/L
RT: 8.069 min Scan# 1014
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion:117 Resp: 306358
Ion Ratio Lower Upper
117 100
119 96.2 73.0 109.4



#33
Benzene
Concen: 27.254 ug/L
RT: 8.325 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 78 Resp: 1042965









#34
Trichloroethene
Concen: 24.693 ug/L
RT: 9.093 min Scan# 1182
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 95 Resp: 240672
Ion Ratio Lower Upper
95 100
97 64.0 44.6 82.8
132 86.6 66.4 123.2
130 93.4 67.3 125.1



#35
Methylcyclohexane
Concen: 13.789 ug/L
RT: 9.331 min Scan# 1221
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 83 Resp: 229916
Ion Ratio Lower Upper
83 100
55 86.4 69.3 103.9
98 46.9 37.4 56.0





#37

1,2-Dichloropropane

Concen: 27.165 ug/L

RT: 9.367 min Scan# 11

Delta R.T. -0.000 min

Lab File: VW026690.D

Acq: 10 Aug 2023 20:32

Instrument :

MSVOA_W

ClientSampleId :





A48K2MS

Tgt Ion: 63 Resp: 284741

Ion Ratio Lower Upper

63 100

112 4.1 3.1 4.7



#38

Bromodichloromethane

Concen: 25.899 ug/L

RT: 9.642 min Scan# 1272

Delta R.T. -0.000 min

Lab File: VW026690.D

Acq: 10 Aug 2023 20:32




Tgt Ion: 83 Resp: 318466


Ion Ratio Lower Upper

83 100

85 63.8 45.4 84.2

127 8.3 6.2 9.2









#39
cis-1,3-Dichloropropene
Concen: 19.209 ug/L
RT: 10.075 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 75 Resp: 289654
Ion Ratio Lower Upper
75 100
77 31.7 22.8 42.4



#40
4-Methyl-2-pentanone
Concen: 58.646 ug/L
RT: 10.209 min Scan# 1365
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 43 Resp: 399860
Ion Ratio Lower Upper
43 100
58 37.1 29.2 43.8
100 12.9 10.7 16.1









#42
Toluene
Concen: 25.851 ug/L
RT: 10.386 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 91 Resp: 1000724
Ion Ratio Lower Upper
91 100
92 57.4 40.8 75.8



#44
trans-1,3-Dichloropropene
Concen: 19.788 ug/L
RT: 10.605 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 75 Resp: 256829
Ion Ratio Lower Upper
75 100
77 32.8 21.2 39.4









#45
1,1,2-Trichloroethane
Concen: 27.043 ug/L
RT: 10.788 min Scan# 1459
Delta R.T. 0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Instrument :
MSVOA_W
ClientSampleId :
A48K2MS

Tgt Ion: 97 Resp: 193867




Ion	Ratio	Lower	Upper
97	100		
99	62.4	44.2	82.0
83	86.0	61.9	114.9
85	55.1	40.0	74.2




#46
Tetrachloroethene
Concen: 19.955 ug/L
RT: 10.855 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 164 Resp: 141300

Ion	Ratio	Lower	Upper
164	100		
129	97.1	68.1	126.5
131	91.1	65.5	121.7
166	129.8	89.1	165.5









#48
2-Hexanone
Concen: 54.633 ug/L
RT: 10.965 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Instrument :
MSVOA_W
ClientSampleId :
A48K2MS

Tgt Ion: 43 Resp: 253467




Ion	Ratio	Lower	Upper
43	100		
58	47.5	41.1	61.7
57	16.8	14.3	21.5
100	9.6	8.8	13.2




#49
Dibromochloromethane
Concen: 24.441 ug/L
RT: 11.123 min Scan# 1515
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 129 Resp: 183253

Ion	Ratio	Lower	Upper
129	100		
127	78.5	56.6	105.2









#50
1,2-Dibromoethane
Concen: 23.654 ug/L
RT: 11.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion:107 Resp: 159105
Ion Ratio Lower Upper
107 100
109 96.9 68.8 127.8
188 3.8 2.7 4.1



#51
Chlorobenzene
Concen: 21.448 ug/L
RT: 11.654 min Scan# 1602
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion:112 Resp: 524462
Ion Ratio Lower Upper
112 100
114 31.0 22.4 41.6
77 63.6 50.2 75.2









#52
Ethylbenzene
Concen: 21.615 ug/L
RT: 11.727 min Scan# 1614
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 91 Resp: 938355
Ion Ratio Lower Upper
91 100
106 28.2 20.2 37.6



#53
m,p-Xylene
Concen: 21.107 ug/L
RT: 11.836 min Scan# 1632
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion:106 Resp: 339452
Ion Ratio Lower Upper
106 100
91 206.8 152.9 284.1









#54
o-Xylene
Concen: 21.134 ug/L
RT: 12.166 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion:106 Resp: 317210
Ion Ratio Lower Upper
106 100
91 215.6 149.9 278.3



#55
Styrene
Concen: 17.586 ug/L
RT: 12.178 min Scan# 1688
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion:104 Resp: 458971
Ion Ratio Lower Upper
104 100
78 48.3 33.0 61.2









#57
1,1,2,2-Tetrachloroethane
Concen: 22.938 ug/L
RT: 12.708 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion: 83 Resp: 208249
Ion Ratio Lower Upper
83 100
85 62.1 44.9 83.3
131 8.6 7.4 11.0



#59
Bromoform
Concen: 35.801 ug/L
RT: 12.349 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion: 173 Resp: 91745
Ion Ratio Lower Upper
173 100
175 48.4 36.6 55.0
254 9.6 7.8 11.6









#60
Isopropylbenzene
Concen: 29.735 ug/L
RT: 12.458 min Scan# 1
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion:	105	Resp:	723205
Ion	Ratio	Lower	Upper
105	100		
120	25.4	21.0	31.6
77	16.8	13.4	20.2



#61
1,2,3-Trichloropropane
Concen: 42.822 ug/L
RT: 12.763 min Scan# 1784
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion:	75	Resp:	169927
Ion	Ratio	Lower	Upper
75	100		
77	32.1	26.1	39.1
110	35.3	29.9	44.9









#62
1,3,5-Trimethylbenzene
Concen: 25.669 ug/L
RT: 12.940 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion:105 Resp: 493461
Ion Ratio Lower Upper
105 100
120 46.0 37.4 56.2



#63
1,2,4-Trimethylbenzene
Concen: 25.243 ug/L
RT: 13.245 min Scan# 1863
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion:105 Resp: 475538
Ion Ratio Lower Upper
105 100
120 44.6 36.2 54.4









#64
1,3-Dichlorobenzene
Concen: 22.287 ug/L
RT: 13.495 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion:146 Resp: 235745
Ion Ratio Lower Upper
146 100
111 46.2 29.1 54.1
148 59.0 38.8 72.0



#65
1,4-Dichlorobenzene
Concen: 21.963 ug/L
RT: 13.574 min Scan# 1917
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion:146 Resp: 242440
Ion Ratio Lower Upper
146 100
111 43.3 28.3 52.7
148 64.4 44.2 82.2









#67
1,2-Dichlorobenzene
Concen: 21.589 ug/L
RT: 13.867 min Scan# 1965
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32




Instrument :
MSVOA_W
ClientSampleId :
A48K2MS


Tgt Ion:	146	Resp:	207661
Ion	Ratio	Lower	Upper
146	100		
111	44.9	34.6	52.0
148	64.3	53.4	80.0



#68
1,2-Dibromo-3-chloropropane
Concen: 31.556 ug/L
RT: 14.476 min Scan# 2065
Delta R.T. -0.006 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32


Tgt Ion:	75	Resp:	30188
Ion	Ratio	Lower	Upper
75	100		
155	69.0	55.0	82.6
157	80.6	77.8	116.6








#69
1,3,5-Trichlorobenzene
Concen: 10.327 ug/L
RT: 14.623 min Scan# 2089
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Instrument :
MSVOA_W
ClientSampleId :
A48K2MS



Tgt Ion:180 Resp: 73800




Ion	Ratio	Lower	Upper
180	100		
182	96.7	75.2	112.8
184	31.6	23.8	35.8
145	32.9	25.1	37.7




#70
1,2,4-trichlorobenzene
Concen: 8.500 ug/L
RT: 15.129 min Scan# 2172
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion:180 Resp: 51503

Ion	Ratio	Lower	Upper
180	100		
182	96.1	72.5	108.7
145	33.4	24.7	37.1









#71
Naphthalene
Concen: 11.396 ug/L
RT: 15.360 min Scan# 211
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Instrument :
MSVOA_W
ClientSampleId :
A48K2MS

Tgt Ion:128 Resp: 131715
Ion Ratio Lower Upper
128 100
127 13.9 10.7 16.1
129 11.2 9.0 13.4



#72
1,2,3-Trichlorobenzene
Concen: 7.596 ug/L
RT: 15.549 min Scan# 2241
Delta R.T. -0.000 min
Lab File: VW026690.D
Acq: 10 Aug 2023 20:32

Tgt Ion:180 Resp: 41510
Ion Ratio Lower Upper
180 100
182 91.6 74.6 112.0
145 33.7 27.0 40.6

