

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081119\
 Data File : VW011808.D
 Acq On : 09 Aug 2019 13:03
 Operator : SY/VA
 Sample : K4215-12
 Misc : 4.65G/10ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ETOC3

Quant Time: Aug 10 04:51:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM073019S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 10 03:43:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	115833	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	59610	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	9433	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	52662	32.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	129.16%
7) Chloroethane-d5	2.89	69	39901	33.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	135.84%
10) 1,1-Dichloroethene-d2	4.01	63	78942	21.11	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	84.44%
20) 2-Butanone-d5	7.07	46	49163	66.94	ug/L	-0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	133.88%
24) Chloroform-d	7.65	84	98756	31.04	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	124.16%
26) 1,2-Dichloroethane-d4	8.31	65	66511	35.33	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	141.32%#
29) Benzene-d6	8.27	84	175611	46.85	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	187.40%#
33) 1,2-Dichloropropane-d6	9.27	67	65423	54.07	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	216.28%#
37) Toluene-d8	10.32	98	102059	30.68	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	122.72%
38) trans-1,3-Dichloropropene-	10.58	79	18380	34.66	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	138.64%#
39) 2-Hexanone-d5	10.93	63	26257	92.90	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	185.80%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	43176	42.96	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	171.84%#
61) 1,2-Dichlorobenzene-d4	13.85	152	10858	29.87	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	119.48%

Target Compounds					Ovalue
13) Acetone	4.12	43	16564	21.058 ug/L	97
16) Methylene chloride	4.92	84	4272	2.363 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081119\
Data File : VW011808.D
Acq On : 09 Aug 2019 13:03
Operator : SY/VA
Sample : K4215-12
Misc : 4.65G/10ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ETOC3

Quant Time: Aug 10 04:51:57 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM073019S.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 10 03:43:50 2019
Response via : Initial Calibration

