

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081120\
 Data File : VW016137.D
 Acq On : 11 Aug 2020 15:36
 Operator : SY/VA
 Sample : L3599-06MS
 Misc : 6.20G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAY89

Quant Time: Aug 11 18:45:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081020S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 11 16:06:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	291058	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	212458	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	48530	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	87665	24.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.16%
7) Chloroethane-d5	2.89	69	71740	26.19	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.76%
10) 1,1-Dichloroethene-d2	4.01	63	193358	25.23	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	100.92%
20) 2-Butanone-d5	7.09	46	73090	76.97	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	153.94%#
24) Chloroform-d	7.65	84	199099	25.49	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	101.96%
26) 1,2-Dichloroethane-d4	8.31	65	116672	28.12	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	112.48%
29) Benzene-d6	8.27	84	376255	32.55	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	130.20%
33) 1,2-Dichloropropane-d6	9.27	67	113462	33.38	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	133.52%#
37) Toluene-d8	10.32	98	295093	27.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	111.36%
38) trans-1,3-Dichloropropene-	10.58	79	42971	28.13	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	112.52%
39) 2-Hexanone-d5	10.93	63	49340	84.15	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	168.30%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	94875	34.29	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	137.16%#
61) 1,2-Dichlorobenzene-d4	13.85	152	49765	28.93	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	115.72%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
12) 1,1-Dichloroethene	4.04	96	90158	24.135	ug/L	95
13) Acetone	4.14	43	14106	19.638	ug/L	98
14) Carbon disulfide	4.37	76	7767	0.662	ug/L	96
34) Benzene	8.32	78	404220	32.993	ug/L	100
35) Trichloroethene	9.09	95	95278	28.921	ug/L	99
44) Toluene	10.39	91	375045	28.999	ug/L	100
52) Chlorobenzene	11.66	112	201892	23.593	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081120\
Data File : VW016137.D
Acq On : 11 Aug 2020 15:36
Operator : SY/VA
Sample : L3599-06MS
Misc : 6.20G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAY89

Quant Time: Aug 11 18:45:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081020S.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 11 16:06:11 2020
Response via : Initial Calibration

