

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081121\
 Data File : VW019710.D
 Acq On : 11 Aug 2021 13:35
 Operator : SY/VA
 Sample : VIBLK401
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK401

Quant Time: Aug 11 18:16:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081121SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Aug 11 18:12:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	348224	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	324691	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	160757	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	123437	30.356	ug/L	0.00
7) Chloroethane-d5	2.897	69	115993	28.860	ug/L	0.00
11) 1,1-Dichloroethene-d2	4.025	63	165682	19.802	ug/L	-0.01
21) 2-Butanone-d5	7.080	46	66841	50.332	ug/L	0.00
24) Chloroform-d	7.653	84	243316	23.112	ug/L	0.00
26) 1,2-Dichloroethane-d4	8.305	65	138174	22.546	ug/L	0.00
32) Benzene-d6	8.281	84	485603	22.588	ug/L	0.00
36) 1,2-Dichloropropane-d6	9.274	67	145905	22.560	ug/L	0.00
41) Toluene-d8	10.323	98	436845	22.276	ug/L	0.00
43) trans-1,3-Dichloroprop...	10.579	79	60702	22.732	ug/L	0.00
47) 2-Hexanone-d5	10.926	63	52123	45.940	ug/L	0.00
56) 1,1,2,2-Tetrachloroeth...	12.694	84	120456	21.465	ug/L	0.00
66) 1,2-Dichlorobenzene-d4	13.853	152	156357	26.371	ug/L	0.00
Target Compounds						
16) Methylene chloride	4.928	84	9619	1.653	ug/L	86
42) Toluene	10.390	91	8492	0.359	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081121\
Data File : VW019710.D
Acq On : 11 Aug 2021 13:35
Operator : SY/VA
Sample : VIBLK401
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK401

Quant Time: Aug 11 18:16:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081121SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 11 18:12:43 2021
Response via : Initial Calibration

