

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081122\
 Data File : VW024225.D
 Acq On : 11 Aug 2022 10:12
 Operator : SY/VA
 Sample : VW0811SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK574

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/12/2022
 Supervised By :Semsettin Yesilyurt 08/12/2022

Quant Time: Aug 12 03:13:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 12 03:13:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	286652	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	288561	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	135832	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	123116	20.548	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	82.200%		
7) Chloroethane-d5	2.873	69	74166	19.870	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	79.480%		
11) 1,1-Dichloroethene-d2	4.001	63	93267	13.658	ug/L	-0.01
Spiked Amount 25.000	Range 45 - 110		Recovery =	54.640%		
21) 2-Butanone-d5	7.074	46	58127	52.144	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	104.280%		
24) Chloroform-d	7.647	84	197418	22.761	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	91.040%		
26) 1,2-Dichloroethane-d4	8.305	65	112300m	23.321	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	93.280%		
32) Benzene-d6	8.269	84	351587	20.659	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	82.640%		
36) 1,2-Dichloropropane-d6	9.268	67	110214	21.935	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	87.720%		
41) Toluene-d8	10.323	98	321019	20.181	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	80.720%		
43) trans-1,3-Dichloroprop...	10.573	79	47241	21.465	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	85.880%		
47) 2-Hexanone-d5	10.927	63	47084	50.732	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	101.460%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84	115473	23.687	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	94.760%		
66) 1,2-Dichlorobenzene-d4	13.853	152	128425	25.837	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	103.360%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081122\
Data File : VW024225.D
Acq On : 11 Aug 2022 10:12
Operator : SY/VA
Sample : VW0811SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK574

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/12/2022
Supervised By :Semsettin Yesilyurt 08/12/2022

Quant Time: Aug 12 03:13:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM080422SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 12 03:13:23 2022
Response via : Initial Calibration

