

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081219\  
 Data File : VW011839.D  
 Acq On : 12 Aug 2019 09:19  
 Operator : SY/VA  
 Sample : VSTDCCC050  
 Misc : 5.00G/5ML/MSVOA W/SOIL  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 LabSampleId :  
 VSTDCCC050

Quant Time: Aug 13 07:54:04 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W072719S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jul 27 00:23:30 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvqRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0    | 94    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.279 | 0.268 | 3.9    | 87    | 0.00     |
| 3 P   | Chloromethane               | 0.483 | 0.404 | 16.4   | 80    | -0.01    |
| 4 C   | Vinyl Chloride              | 0.582 | 0.559 | 4.0#   | 86    | 0.00     |
| 5 T   | Bromomethane                | 0.322 | 0.299 | 7.1    | 86    | 0.00     |
| 6 T   | Chloroethane                | 0.332 | 0.314 | 5.4    | 83    | 0.00     |
| 7 T   | Trichlorofluoromethane      | 0.260 | 0.317 | -21.9# | 105   | -0.01    |
| 8 T   | Diethyl Ether               | 0.295 | 0.266 | 9.8    | 79    | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.493 | 0.506 | -2.6   | 90    | -0.01    |
| 10 T  | Methyl Iodide               | 0.697 | 0.687 | 1.4    | 84    | -0.01    |
| 11 T  | Tert butyl alcohol          | 0.048 | 0.018 | 62.5#  | 38#   | 0.13     |
| 12 CM | 1,1-Dichloroethene          | 0.521 | 0.512 | 1.7#   | 86    | 0.00     |
| 13 T  | Acrolein                    | 0.041 | 0.033 | 19.5   | 77    | 0.02     |
| 14 T  | Allyl chloride              | 1.108 | 1.112 | -0.4   | 88    | 0.00     |
| 15 T  | Acrylonitrile               | 0.147 | 0.130 | 11.6   | 77    | 0.02     |
| 16 T  | Acetone                     | 0.183 | 0.193 | -5.5   | 93    | 0.06     |
| 17 T  | Carbon Disulfide            | 1.630 | 1.568 | 3.8    | 83    | -0.01    |
| 18 T  | Methyl Acetate              | 0.410 | 0.342 | 16.6   | 72    | 0.02     |
| 19 T  | Methyl tert-butyl Ether     | 0.826 | 0.776 | 6.1    | 80    | 0.00     |
| 20 T  | Methylene Chloride          | 0.630 | 0.546 | 13.3   | 82    | -0.01    |
| 21 T  | trans-1,2-Dichloroethene    | 0.561 | 0.547 | 2.5    | 87    | 0.00     |
| 22 T  | Diisopropyl ether           | 2.088 | 1.983 | 5.0    | 83    | 0.00     |
| 23 T  | Vinyl Acetate               | 1.312 | 1.226 | 6.6    | 80    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 1.118 | 1.096 | 2.0    | 86    | 0.00     |
| 25 T  | 2-Butanone                  | 0.239 | 0.219 | 8.4    | 77    | 0.02     |
| 26 T  | 2,2-Dichloropropane         | 0.688 | 0.640 | 7.0    | 88    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.618 | 0.579 | 6.3    | 83    | 0.00     |
| 28 T  | Bromochloromethane          | 0.496 | 0.469 | 5.4    | 91    | 0.00     |
| 29 T  | Tetrahydrofuran             | 0.137 | 0.116 | 15.3   | 73    | 0.01     |
| 30 C  | Chloroform                  | 1.032 | 0.989 | 4.2#   | 86    | 0.00     |
| 31 T  | Cyclohexane                 | 1.073 | 1.007 | 6.2    | 86    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 0.786 | 0.803 | -2.2   | 89    | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.574 | 0.535 | 6.8    | 85    | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0    | 91    | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.281 | 0.277 | 1.4    | 85    | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.520 | 0.539 | -3.7   | 88    | 0.00     |
| 37 T  | Ethyl Acetate               | 0.299 | 0.269 | 10.0   | 75    | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.445 | 0.475 | -6.7   | 90    | 0.00     |
| 39 T  | Methylcyclohexane           | 0.572 | 0.578 | -1.0   | 85    | 0.00     |
| 40 TM | Benzene                     | 1.456 | 1.451 | 0.3    | 86    | 0.00     |
| 41 T  | Methacrylonitrile           | 0.179 | 0.160 | 10.6   | 76    | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 0.456 | 0.444 | 2.6    | 83    | 0.00     |
| 43 T  | Isopropyl Acetate           | 0.550 | 0.503 | 8.5    | 76    | 0.00     |
| 44 TM | Trichloroethene             | 0.350 | 0.353 | -0.9   | 86    | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.390 | 0.384 | 1.5#   | 83    | 0.00     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081219\  
 Data File : VW011839.D  
 Acq On : 12 Aug 2019 09:19  
 Operator : SY/VA  
 Sample : VSTDCCC050  
 Misc : 5.00G/5ML/MSVOA W/SOIL  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 LabSampleId :  
 VSTDCCC050

Quant Time: Aug 13 07:54:04 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W072719S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Jul 27 00:23:30 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvqRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 46 T  | Dibromomethane              | 0.178 | 0.173 | 2.8   | 82    | 0.00     |
| 47 T  | Bromodichloromethane        | 0.467 | 0.469 | -0.4  | 85    | 0.00     |
| 48 T  | Methyl methacrylate         | 0.264 | 0.238 | 9.8   | 73    | 0.00     |
| 49 T  | 1,4-Dioxane                 | 0.003 | 0.001 | 66.7# | 36#   | 0.07     |
| 50 S  | Toluene-d8                  | 1.148 | 1.134 | 1.2   | 86    | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.286 | 0.260 | 9.1   | 74    | 0.00     |
| 52 CM | Toluene                     | 0.879 | 0.877 | 0.2#  | 85    | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.487 | 0.475 | 2.5   | 81    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.579 | 0.571 | 1.4   | 82    | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.252 | 0.241 | 4.4   | 80    | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.369 | 0.350 | 5.1   | 76    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.477 | 0.456 | 4.4   | 80    | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.168 | 0.191 | -13.7 | 92    | 0.00     |
| 59 T  | 2-Hexanone                  | 0.202 | 0.214 | -5.9  | 84    | 0.00     |
| 60 T  | Dibromochloromethane        | 0.280 | 0.275 | 1.8   | 82    | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.239 | 0.225 | 5.9   | 79    | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.409 | 0.393 | 3.9   | 83    | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0   | 88    | 0.00     |
| 64 T  | Tetrachloroethene           | 0.326 | 0.329 | -0.9  | 86    | 0.00     |
| 65 PM | Chlorobenzene               | 1.005 | 0.998 | 0.7   | 83    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.350 | 0.356 | -1.7  | 83    | 0.00     |
| 67 C  | Ethyl Benzene               | 1.915 | 1.967 | -2.7# | 84    | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.696 | 0.712 | -2.3  | 84    | 0.00     |
| 69 T  | o-Xylene                    | 0.647 | 0.654 | -1.1  | 82    | 0.00     |
| 70 T  | Styrene                     | 1.119 | 1.139 | -1.8  | 83    | 0.00     |
| 71 P  | Bromoform                   | 0.179 | 0.178 | 0.6   | 78    | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0   | 89    | 0.00     |
| 73 T  | Isopropylbenzene            | 3.730 | 3.808 | -2.1  | 84    | 0.00     |
| 74 T  | N-amyl acetate              | 1.168 | 1.102 | 5.7   | 74    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 0.702 | 0.667 | 5.0   | 77    | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 0.529 | 0.473 | 10.6  | 78    | 0.00     |
| 77 T  | Bromobenzene                | 0.824 | 0.820 | 0.5   | 82    | 0.00     |
| 78 T  | n-propylbenzene             | 4.540 | 4.662 | -2.7  | 85    | 0.00     |
| 79 T  | 2-Chlorotoluene             | 2.615 | 2.617 | -0.1  | 84    | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 3.127 | 3.191 | -2.0  | 84    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.230 | 0.227 | 1.3   | 77    | 0.00     |
| 82 T  | 4-Chlorotoluene             | 2.744 | 2.708 | 1.3   | 82    | 0.00     |
| 83 T  | tert-Butylbenzene           | 2.615 | 2.668 | -2.0  | 83    | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 3.150 | 3.159 | -0.3  | 83    | 0.00     |
| 85 T  | sec-Butylbenzene            | 3.731 | 3.770 | -1.0  | 83    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.385 | 3.396 | -0.3  | 82    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.617 | 1.596 | 1.3   | 82    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 1.608 | 1.579 | 1.8   | 83    | 0.00     |
| 89 T  | n-Butylbenzene              | 3.378 | 3.378 | 0.0   | 82    | 0.00     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081219\  
Data File : VW011839.D  
Acq On : 12 Aug 2019 09:19  
Operator : SY/VA  
Sample : VSTDCCC050  
Misc : 5.00G/5ML/MSVOA W/SOIL  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
LabSampleId :  
VSTDCCC050

Quant Time: Aug 13 07:54:04 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W072719S.M  
Quant Title : SW846 8260  
QLast Update : Sat Jul 27 00:23:30 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 90 T | Hexachloroethane            | 0.599 | 0.617 | -3.0 | 84    | 0.00     |
| 91 T | 1,2-Dichlorobenzene         | 1.430 | 1.382 | 3.4  | 80    | 0.00     |
| 92 T | 1,2-Dibromo-3-Chloropropane | 0.120 | 0.113 | 5.8  | 75    | 0.00     |
| 93 T | 1,2,4-Trichlorobenzene      | 1.008 | 0.969 | 3.9  | 78    | 0.00     |
| 94 T | Hexachlorobutadiene         | 0.628 | 0.620 | 1.3  | 82    | 0.00     |
| 95 T | Naphthalene                 | 1.787 | 1.755 | 1.8  | 75    | 0.00     |
| 96 T | 1,2,3-Trichlorobenzene      | 0.883 | 0.847 | 4.1  | 78    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 6