

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW081318\
 Data File : VW004600.D
 Acq On : 10 Aug 2018 16:32
 Operator : SY/AP
 Sample : J4409-01
 Misc : 5.02G/5ML/MSVOA_W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-9B

Quant Time: Aug 11 01:01:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 08:22:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	663723	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	1157183	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	1009597	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	416004	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	370186	50.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.76%	
35) Dibromofluoromethane	7.88	113	340236	44.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.20%	
50) Toluene-d8	10.33	98	1176717	40.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.00%	
62) 4-Bromofluorobenzene	12.62	95	386375	36.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	72.92%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	5.19	59	1139	2.648	ug/l #	70
16) Acetone	4.13	43	80803	44.038	ug/l	98
20) Methylene Chloride	4.91	84	49620	Below Cal		95
25) 2-Butanone	7.18	43	15592	5.283	ug/l #	86
31) Cyclohexane	7.95	56	17212	1.251	ug/l #	1
36) 1,1-Dichloropropene	7.95	75	52398	4.205	ug/l #	49
37) Ethyl Acetate	7.18	43	15592	2.672	ug/l #	69
38) Carbon Tetrachloride	7.95	117	62975	5.617	ug/l #	17
52) Toluene	10.39	92	23000	1.066	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.90	63	37	8.952	ug/l #	48
59) 2-Hexanone	10.94	43	6768	1.410	ug/l #	30
67) Ethyl Benzene	11.74	91	192810	5.112	ug/l	99
68) m/p-Xylenes	11.85	106	20084	1.378	ug/l	97
73) Isopropylbenzene	12.47	105	112410	3.907	ug/l	99
74) N-amyl acetate	12.26	43	9970	1.429	ug/l #	66
76) 1,2,3-Trichloropropane	12.62	75	188732	43.843	ug/l #	100
78) n-propylbenzene	12.81	91	84147	2.370	ug/l	99
79) 2-Chlorotoluene	12.91	91	67182	3.270	ug/l #	41
80) 1,3,5-Trimethylbenzene	12.95	105	192317	7.835	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.62	75	188732	104.954	ug/l #	13
83) tert-Butylbenzene	13.26	119	57879	2.811	ug/l #	67
84) 1,2,4-Trimethylbenzene	13.26	105	435223	17.128	ug/l	100
86) p-Isopropyltoluene	13.51	119	184112	6.841	ug/l	98
89) n-Butylbenzene	13.80	91	320433	12.117	ug/l #	1
90) Hexachloroethane	14.10	117	109843	22.058	ug/l #	9
92) 1,2-Dibromo-3-Chloropropan	14.47	75	2783	2.835	ug/l #	2
95) Naphthalene	15.38	128	11648304	592.821	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW081318\
Data File : VW004600.D
Acq On : 10 Aug 2018 16:32
Operator : SY/AP
Sample : J4409-01
Misc : 5.02G/5ML/MSVOA_W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-CLARK-AREA-9B

Quant Time: Aug 11 01:01:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 08:22:01 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 1044

Delta R.T. 0.00 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32

Instrument :

MSVOA_W

ClientSampleId :


2018-NYSEG-CLARK-AREA-9B

Tgt Ion:168 Resp: 663723

Ion Ratio Lower Upper

168 100

99 54.0 44.1 66.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95



300000

200000

100000

0

Time-->



#11

Tert butyl alcohol

Concen: 2.648 ug/l

RT: 5.19 min Scan# 591

Delta R.T. -0.02 min

Lab File: VW004600.D


Acq: 10 Aug 2018 16:32

Tgt Ion: 59 Resp: 1139

Ion Ratio Lower Upper

59 100

57 0.0 9.1 13.7#



Abundance Ion 59.00 (58.70 to 59.70): VW

Ion 57.00 (56.70 to 57.70): VW

5.19

400


300


200

100

0

Time-->









#16
Acetone
Concen: 44.038 ug/l
RT: 4.13 min Scan# 417
Delta R.T. -0.01 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32




Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-CLARK-AREA-9B


Tgt Ion: 43 Resp: 80803
Ion Ratio Lower Upper
43 100
58 32.4 26.8 40.2



#20
Methylene Chloride
Concen: Below Cal
RT: 4.91 min Scan# 545
Delta R.T. 0.00 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32

Tgt Ion: 84 Resp: 49620
Ion Ratio Lower Upper
84 100
49 113.8 99.6 149.4
51 38.7 31.0 46.4
86 64.6 51.5 77.3









#25
2-Butanone
Concen: 5.283 ug/l
RT: 7.18 min Scan# 917
Delta R.T. -0.01 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32




Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-CLARK-AREA-9B


Tgt Ion: 43 Resp: 15592
Ion Ratio Lower Upper
43 100
72 19.2 21.0 31.6#



#31
Cyclohexane
Concen: 1.251 ug/l
RT: 7.95 min Scan# 1043
Delta R.T. -0.01 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32

Tgt Ion: 56 Resp: 17212
Ion Ratio Lower Upper
56 100
69 177.5 26.4 39.6#
84 153.3 66.9 100.3#









#33
 1,2-Dichloroethane-d4
 Concen: 50.385 ug/l
 RT: 8.31 min Scan# 1103
 Delta R.T. 0.00 min
 Lab File: VW004600.D
 Acq: 10 Aug 2018 16:32




Instrument :
 MSVOA_W
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-9B


Tgt Ion: 65 Resp: 370186
 Ion Ratio Lower Upper
 65 100
 67 54.0 0.0 109.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.85 min Scan# 1191
 Delta R.T. 0.00 min
 Lab File: VW004600.D
 Acq: 10 Aug 2018 16:32

Tgt Ion: 114 Resp: 1157183
 Ion Ratio Lower Upper
 114 100
 63 20.3 0.0 42.4
 88 15.4 0.0 31.2









#35
 Dibromofluoromethane
 Concen: 44.096 ug/l
 RT: 7.88 min Scan# 1033
 Delta R.T. 0.00 min
 Lab File: VW004600.D
 Acq: 10 Aug 2018 16:32




Instrument :
 MSVOA_W
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-9B


Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.1	82.9	124.3
192	20.5	15.8	23.6



#36
 1,1-Dichloropropene
 Concen: 4.205 ug/l
 RT: 7.95 min Scan# 1043
 Delta R.T. -0.13 min
 Lab File: VW004600.D
 Acq: 10 Aug 2018 16:32

Tgt Ion	Ratio	Lower	Upper
75	100		
110	7.6	17.4	52.2#
77	0.5	24.7	37.1#









#37
Ethyl Acetate
Concen: 2.672 ug/l
RT: 7.18 min Scan# 917
Delta R.T. -0.08 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32




Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-CLARK-AREA-9B


Tgt Ion: 43 Resp: 15592
Ion Ratio Lower Upper
43 100
61 0.0 11.9 17.9#
70 1.3 8.2 12.4#



#38
Carbon Tetrachloride
Concen: 5.617 ug/l
RT: 7.95 min Scan# 1044
Delta R.T. -0.12 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32

Tgt Ion: 117 Resp: 62975
Ion Ratio Lower Upper
117 100
119 5.8 76.1 114.1#
121 0.0 24.2 36.4#





#50

Toluene-d8

Concen: 40.004 ug/l

RT: 10.33 min Scan# 1434

Delta R.T. 0.00 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32

Instrument :

MSVOA_W

ClientSampleId :





2018-NYSEG-CLARK-AREA-9B

Tgt Ion: 98 Resp: 1176717

Ion Ratio Lower Upper

98 100

100 63.6 51.4 77.2



#52

Toluene

Concen: 1.066 ug/l

RT: 10.39 min Scan# 1444

Delta R.T. 0.00 min

Lab File: VW004600.D




Acq: 10 Aug 2018 16:32


Tgt Ion: 92 Resp: 23000

Ion Ratio Lower Upper

92 100

91 174.5 137.9 206.9





#58

2-Chloroethyl Vinyl ether

Concen: 8.952 ug/l

RT: 9.90 min Scan# 1364

Delta R.T. -0.02 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32

Instrument :

MSVOA_W

ClientSampleId :


2018-NYSEG-CLARK-AREA-9B

Tgt Ion: 63 Resp: 37

Ion Ratio Lower Upper

63 100

106 0.0 21.3 31.9#



Abundance Ion 62.90 (62.60 to 63.60): VW

Ion 105.90 (105.60 to 106.60): VW

9.90

100

80

60



40

20

0

9.88 9.89 9.90 9.91 9.92

Time-->



#59

2-Hexanone

Concen: 1.410 ug/l

RT: 10.94 min Scan# 1534

Delta R.T. -0.04 min

Lab File: VW004600.D


Acq: 10 Aug 2018 16:32

Tgt Ion: 43 Resp: 6768

Ion Ratio Lower Upper

43 100

58 4.5 27.9 83.6#



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW

10.94

4000

3000


2000


1000

0

10.85 10.90 10.95 11.00

Time-->





#62

4-Bromofluorobenzene

Concen: 36.465 ug/l

RT: 12.62 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32

Instrument :

MSVOA_W

ClientSampleId :

2018-NYSEG-CLARK-AREA-9B





Tgt Ion: 95 Resp: 386375

Ion Ratio Lower Upper

95 100

174 80.1 0.0 160.4

176 76.5 0.0 153.2



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.63 min Scan# 1648

Delta R.T. 0.00 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32




Tgt Ion: 117 Resp: 1009597


Ion Ratio Lower Upper

117 100

82 57.1 45.3 67.9

119 31.8 26.2 39.4









#67
Ethyl Benzene
Concen: 5.112 ug/l
RT: 11.74 min Scan# 1665
Delta R.T. 0.00 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32




Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-CLARK-AREA-9B


Tgt Ion: 91 Resp: 192810
Ion Ratio Lower Upper
91 100
106 29.8 24.2 36.2



#68
m/p-Xylenes
Concen: 1.378 ug/l
RT: 11.85 min Scan# 1683
Delta R.T. 0.00 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32

Tgt Ion: 106 Resp: 20084
Ion Ratio Lower Upper
106 100
91 207.6 162.8 244.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.57 min Scan# 1965

Delta R.T. 0.00 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32

Instrument :

MSVOA_W

ClientSampleId :

2018-NYSEG-CLARK-AREA-9B


Tgt Ion:152 Resp: 416004

Ion Ratio Lower Upper

152 100

115 58.0 29.1 87.3




150 156.1 0.0 349.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 3.907 ug/l

RT: 12.47 min Scan# 1785

Delta R.T. 0.00 min

Lab File: VW004600.D


Acq: 10 Aug 2018 16:32

Tgt Ion:105 Resp: 112410

Ion Ratio Lower Upper



105 100


120 26.1 13.2 39.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#74

N-ethyl acetate

Concen: 1.429 ug/l

RT: 12.26 min Scan# 1751

Delta R.T. -0.02 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32

Instrument :

MSVOA_W

ClientSampleId :

2018-NYSEG-CLARK-AREA-9B

Tgt Ion: 43 Resp: 9970





Ion Ratio Lower Upper

43 100

70 34.4 35.3 52.9#

55 55.0 22.9 34.3#

61 0.3 19.7 29.5#



#76

1,2,3-Trichloropropane

Concen: 43.843 ug/l

RT: 12.62 min Scan# 1810

Delta R.T. -0.15 min

Lab File: VW004600.D




Acq: 10 Aug 2018 16:32


Tgt Ion: 75 Resp: 188732

Ion Ratio Lower Upper

75 100

77 2.4 0.0 0.0#









#78
n-propylbenzene
Concen: 2.370 ug/l
RT: 12.81 min Scan# 1841
Delta R.T. 0.00 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32




Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-CLARK-AREA-9B


Tgt Ion: 91 Resp: 84147
Ion Ratio Lower Upper
91 100
120 23.3 11.3 33.9



#79
2-Chlorotoluene
Concen: 3.270 ug/l
RT: 12.91 min Scan# 1857
Delta R.T. 0.01 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32

Tgt Ion: 91 Resp: 67182
Ion Ratio Lower Upper
91 100
126 0.0 16.6 49.8#









#80
 1,3,5-Trimethylbenzene
 Concen: 7.835 ug/l
 RT: 12.95 min Scan# 1864
 Delta R.T. 0.00 min
 Lab File: VW004600.D
 Acq: 10 Aug 2018 16:32




Instrument :
 MSVOA_W
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-9B


Tgt Ion:105 Resp: 192317
 Ion Ratio Lower Upper
 105 100
 120 48.5 24.6 73.8



#81
 trans-1,4-Dichloro-2-butene
 Concen: 104.954 ug/l
 RT: 12.62 min Scan# 1810
 Delta R.T. 0.10 min
 Lab File: VW004600.D
 Acq: 10 Aug 2018 16:32

Tgt Ion: 75 Resp: 188732
 Ion Ratio Lower Upper
 75 100
 53 0.5 74.4 111.6#
 89 0.1 35.9 53.9#









#83
 tert-Butylbenzene
 Concen: 2.811 ug/l
 RT: 13.26 min Scan# 1914
 Delta R.T. 0.04 min
 Lab File: VW004600.D
 Acq: 10 Aug 2018 16:32




Instrument :
 MSVOA_W
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-9B


Tgt Ion:119 Resp: 57879
 Ion Ratio Lower Upper
 119 100
 91 87.8 33.3 99.8
 134 0.0 12.2 36.6#



#84
 1,2,4-Trimethylbenzene
 Concen: 17.128 ug/l
 RT: 13.26 min Scan# 1914
 Delta R.T. 0.00 min
 Lab File: VW004600.D
 Acq: 10 Aug 2018 16:32

Tgt Ion:105 Resp: 435223
 Ion Ratio Lower Upper
 105 100
 120 45.3 22.6 67.8









#86
p-Isopropyltoluene
Concen: 6.841 ug/l
RT: 13.51 min Scan# 1955
Delta R.T. 0.00 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32




Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-CLARK-AREA-9B


Tgt Ion: 119 Resp: 184112
Ion Ratio Lower Upper
119 100
134 26.2 13.1 39.1
91 24.8 11.4 34.1



#89
n-Butylbenzene
Concen: 12.117 ug/l
RT: 13.80 min Scan# 2003
Delta R.T. -0.03 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32

Tgt Ion: 91 Resp: 320433
Ion Ratio Lower Upper
91 100
92 7.2 26.4 79.0#
134 129.5 12.4 37.2#





#90

Hexachloroethane

Concen: 22.058 ug/l

RT: 14.10 min Scan# 2053

Delta R.T. 0.00 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32

Instrument :

MSVOA_W

ClientSampleId :





2018-NYSEG-CLARK-AREA-9B

Tgt Ion:117 Resp: 109843

Ion Ratio Lower Upper

117 100

201 0.0 40.1 120.3#



#92

1,2-Dibromo-3-Chloropropane

Concen: 2.835 ug/l

RT: 14.47 min Scan# 2113

Delta R.T. -0.02 min

Lab File: VW004600.D

Acq: 10 Aug 2018 16:32




Tgt Ion: 75 Resp: 2783


Ion Ratio Lower Upper

75 100

155 0.0 41.3 123.7#

157 0.0 52.6 157.8#





#95
Naphthalene
Concen: 592.821 ug/l
RT: 15.38 min Scan# 2262
Delta R.T. 0.00 min
Lab File: VW004600.D
Acq: 10 Aug 2018 16:32

Instrument :
MSVOA_W
ClientSampleId :
2018-NYSEG-CLARK-AREA-9B

Tgt Ion:128 Resp:11648304
Ion Ratio Lower Upper
128 100
127 13.3 10.3 15.5
129 11.3 8.6 13.0

