

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081318\
 Data File : VW004627.D
 Acq On : 11 Aug 2018 05:52
 Operator : SY/AP
 Sample : J4272-19
 Misc : 578G/5ML/MSVOA W/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-03-080918-J

Quant Time: Aug 11 06:41:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 08:22:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	640178	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	1078956	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	884165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	290371	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	348633	49.20	ug/l	0.00
Spiked Amount			Recovery	=	98.40%	
35) Dibromofluoromethane	7.88	113	318555	44.28	ug/l	0.00
Spiked Amount			Recovery	=	88.56%	
50) Toluene-d8	10.33	98	971323	35.42	ug/l	0.00
Spiked Amount			Recovery	=	70.84%	
62) 4-Bromofluorobenzene	12.62	95	267725	27.10	ug/l	0.00
Spiked Amount			Recovery	=	54.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.14	43	28762	16.25	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081318\
Data File : VW004627.D
Acq On : 11 Aug 2018 05:52
Operator : SY/AP
Sample : J4272-19
Misc : 578G/5ML/MSVOA W/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JC-03-080918-J

Quant Time: Aug 11 06:41:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 08:22:01 2018
Response via : Initial Calibration

