

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081318\
 Data File : VW004609.D
 Acq On : 10 Aug 2018 20:24
 Operator : SY/AP
 Sample : J4430-12
 Misc : 8.56G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 G.L-WC-01

Quant Time: Aug 11 02:45:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 08:22:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	672747	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	1146963	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	924336	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	275383	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	362544	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
35) Dibromofluoromethane	7.88	113	337216	44.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.18%	
50) Toluene-d8	10.33	98	1092013	37.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.90%	
62) 4-Bromofluorobenzene	12.62	95	275306	26.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	52.42%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081318\
Data File : VW004609.D
Acq On : 10 Aug 2018 20:24
Operator : SY/AP
Sample : J4430-12
Misc : 8.56G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
G.L-WC-01

Quant Time: Aug 11 02:45:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 08:22:01 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.95 min Scan# 1044

Delta R.T. 0.00 min

Lab File: VW004609.D

Acq: 10 Aug 2018 20:24

Instrument :

MSVOA_W

ClientSampled :


G.L-WC-01

Tot Ion:168 Resp: 672747

Ion Ratio Lower Upper

168 100

99 52.6 44.1 66.1



Abundance Ion 167.90 (167.60 to 168.60): VW

Ion 98.90 (98.60 to 99.60): VW

7.95



300000

200000

100000

0

Time--> 7.80 7.90 8.00 8.10



#33

1,2-Dichloroethane-d4

Concen: 48.68 ug/l

RT: 8.31 min Scan# 1103

Delta R.T. 0.00 min

Lab File: VW004609.D


Acq: 10 Aug 2018 20:24

Tgt Ion: 65 Resp: 362544

Ion Ratio Lower Upper

65 100

67 54.6 0.0 109.0



Abundance Ion 64.90 (64.60 to 65.60): VW

Ion 66.90 (66.60 to 67.60): VW

8.31


150000


100000

50000

0

Time--> 8.20 8.30 8.40









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.85 min Scan# 1191
 Delta R.T. 0.00 min
 Lab File: VW004609.D
 Acq: 10 Aug 2018 20:24

Instrument :
 MSVOA_W
 ClientSampleId :
 G.L-WC-01

Tot Ion:114 Resp: 1146963
 Ion Ratio Lower Upper
 114 100
 63 19.9 0.0 42.4
 88 16.1 0.0 31.2




Abundance Ion 113.90 (113.60 to 114.60): V
 800000 Ion 63.00 (62.70 to 63.70): VW
 600000 Ion 87.90 (87.60 to 88.60): VW






#35
 Dibromofluoromethane
 Concen: 44.09 ug/l
 RT: 7.88 min Scan# 1033
 Delta R.T. 0.00 min
 Lab File: VW004609.D
 Acq: 10 Aug 2018 20:24

Tgt Ion:113 Resp: 337216
 Ion Ratio Lower Upper
 113 100
 111 103.3 82.9 124.3
 192 20.7 15.8 23.6



Abundance Ion 112.80 (112.50 to 113.50): V
 200000 Ion 110.80 (110.50 to 111.50): V
 150000 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 37.45 ug/l

RT: 10.33 min Scan# 1434

Delta R.T. 0.00 min

Lab File: VW004609.D

Acq: 10 Aug 2018 20:24

Instrument :

MSVOA_W

ClientSampleId :





G.L-WC-01

Tot Ion: 98 Resp: 1092013

Ion Ratio Lower Upper

98 100

100 63.6 51.4 77.2



#62

4-Bromofluorobenzene

Concen: 26.21 ug/l

RT: 12.62 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VW004609.D

Acq: 10 Aug 2018 20:24




Tgt Ion: 95 Resp: 275306


Ion Ratio Lower Upper

95 100

174 81.5 0.0 160.4

176 77.5 0.0 153.2






#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.63 min Scan# 1648
Delta R.T. 0.00 min
Lab File: VW004609.D
Acq: 10 Aug 2018 20:24

Instrument :
MSVOA_W
ClientSampleId :
G.L-WC-01

Tot Ion	Ratio	Lower	Upper
117	100		
82	56.2	45.3	67.9
119	31.5	26.2	39.4






Abundance

Ion 116.90 (116.60 to 117.60): V


Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.57 min Scan# 1965
Delta R.T. 0.00 min
Lab File: VW004609.D
Acq: 10 Aug 2018 20:24

Tgt Ion	Ratio	Lower	Upper
152	100		
115	58.5	29.1	87.3
150	156.9	0.0	349.6



Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

