

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081621\
 Data File : VW019747.D
 Acq On : 16 Aug 2021 12:20
 Operator : SY/VA
 Sample : M3365-14RE
 Misc : 8.93g/10mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBKCORE

Quant Time: Aug 17 01:53:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081121SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Aug 17 01:50:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	23697	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	23007	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	9652	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	8988	32.480	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	129.920%
7) Chloroethane-d5	2.891	69	10305	37.677	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	150.720%#
11) 1,1-Dichloroethene-d2	4.031	63	10422	18.304	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.200%
21) 2-Butanone-d5	7.092	46	5408	59.842	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	119.680%
24) Chloroform-d	7.653	84	19930	27.819	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	111.280%
26) 1,2-Dichloroethane-d4	8.311	65	11226	26.917	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.680%
32) Benzene-d6	8.281	84	31434	20.635	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.560%
36) 1,2-Dichloropropane-d6	9.274	67	11201	24.442	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.760%
41) Toluene-d8	10.329	98	22878	16.464	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	65.840%
43) trans-1,3-Dichloroprop...	10.579	79	3895	20.585	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.320%
47) 2-Hexanone-d5	10.926	63	3014	37.490	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.980%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	10397	26.147	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	9266	26.029	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.120%
Target Compounds					Qvalue	
13) Acetone	4.141	43	1587	21.364	ug/L	96
16) Methylene chloride	4.927	84	3580	9.040	ug/L	84
42) Toluene	10.390	91	510	0.305	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081621\
Data File : VW019747.D
Acq On : 16 Aug 2021 12:20
Operator : SY/VA
Sample : M3365-14RE
Misc : 8.93g/10mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBKCORE

Quant Time: Aug 17 01:53:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081121SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Aug 17 01:50:39 2021
Response via : Initial Calibration

