

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081624\
 Data File : VW029931.D
 Acq On : 16 Aug 2024 15:11
 Operator : SY/MD
 Sample : P3504-18
 Misc : 5.59g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCYF4

Quant Time: Aug 17 00:36:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Aug 17 00:33:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	264863	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	223865	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	77789	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	66395	17.657	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.640%
7) Chloroethane-d5	2.899	69	56934	21.208	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.840%
11) 1,1-Dichloroethene-d2	4.015	65	27331	15.923	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.680%
21) 2-Butanone-d5	7.075	46	33559	32.562	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	65.120%
24) Chloroform-d	7.648	84	133334	19.447	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.800%
26) 1,2-Dichloroethane-d4	8.307	65	78484	20.097	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.400%
32) Benzene-d6	8.276	84	242559	20.053	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.200%
36) 1,2-Dichloropropane-d6	9.270	67	77600	20.092	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.360%
41) Toluene-d8	10.319	98	207241	19.269	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.080%
43) trans-1,3-Dichloroprop...	10.575	79	29004	18.558	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.240%
47) 2-Hexanone-d5	10.922	63	25874	40.628	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.260%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	62780	21.891	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.560%
66) 1,2-Dichlorobenzene-d4	13.848	152	51982	21.521	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.080%
Target Compounds						Qvalue
13) Acetone	4.125	43	22407	17.495	ug/L	92
15) Methyl Acetate	4.679	43	4962	2.326	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081624\
Data File : VW029931.D
Acq On : 16 Aug 2024 15:11
Operator : SY/MD
Sample : P3504-18
Misc : 5.59g/10mL/MSVOA_W/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCYF4

Quant Time: Aug 17 00:36:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072424SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Aug 17 00:33:10 2024
Response via : Initial Calibration

