


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW081718\
 Data File : VW004700.D
 Acq On : 15 Aug 2018 16:43
 Operator : SY/AP
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
 Title : SW846 8260
 Last Update : Tue Aug 07 08:22:01 2018



AutoFind: Scans 1810, 1811, 1812; Background Corrected with Scan 1801

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	55690	PASS
75	95	30	60	47.1	153829	PASS
95	95	100	100	100.0	326421	PASS
96	95	5	9	6.8	22213	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	85.2	278122	PASS
175	174	5	9	7.2	20061	PASS
176	174	95	101	96.4	268096	PASS
177	176	5	9	6.7	18092	PASS