

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081719\
 Data File : VW011955.D
 Acq On : 16 Aug 2019 11:01
 Operator : SY/VA
 Sample : K3616-01
 Misc : 6.94G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-03-081519-A

Quant Time: Aug 17 05:11:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	204203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	345736	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	276384	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	104394	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	133174	56.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.60%	
35) Dibromofluoromethane	7.88	113	105791	54.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.08%	
50) Toluene-d8	10.32	98	412811	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
62) 4-Bromofluorobenzene	12.62	95	125985	44.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.00%	
Target Compounds						
16) Acetone	4.14	43	33932	45.492	ug/l	100
20) Methylene Chloride	4.92	84	31749	12.342	ug/l	96
29) Tetrahydrofuran	7.54	42	6935	12.431	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081719\
Data File : VW011955.D
Acq On : 16 Aug 2019 11:01
Operator : SY/VA
Sample : K3616-01
Misc : 6.94G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JC-03-081519-A

Quant Time: Aug 17 05:11:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:23:30 2019
Response via : Initial Calibration

