

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081821\
 Data File : VW019772.D
 Acq On : 18 Aug 2021 14:28
 Operator : SY/VA
 Sample : M3458-02
 Misc : 7.23g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AB6

Quant Time: Aug 19 05:16:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Aug 19 05:14:25 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	289529	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	232886	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	65556	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.367	65	159158	35.479	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 141.920%		
7) Chloroethane-d5	2.897	69	133864	33.997	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 136.000%		
11) 1,1-Dichloroethene-d2	4.031	63	177525	22.806	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 91.240%		
21) 2-Butanone-d5	7.085	46	46474	51.003	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 102.000%		
24) Chloroform-d	7.652	84	234929	29.505	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 118.000%		
26) 1,2-Dichloroethane-d4	8.311	65	130489	30.410	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 121.640%		
32) Benzene-d6	8.280	84	458753	34.548	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 138.200%#		
36) 1,2-Dichloropropane-d6	9.280	67	140348	34.903	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 139.600%#		
41) Toluene-d8	10.323	98	380721	31.287	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 125.160%		
43) trans-1,3-Dichloroprop...	10.579	79	51319	30.893	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 123.560%		
47) 2-Hexanone-d5	10.926	63	32351	57.751	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 115.500%		
56) 1,1,2,2-Tetrachloroeth...	12.694	84	100352	30.662	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 122.640%#		
66) 1,2-Dichlorobenzene-d4	13.853	152	66912	29.756	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 119.040%		
Target Compounds						
42) Toluene	10.390	91	6641	0.439	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081821\
Data File : VW019772.D
Acq On : 18 Aug 2021 14:28
Operator : SY/VA
Sample : M3458-02
Misc : 7.23g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AB6

Quant Time: Aug 19 05:16:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Aug 19 05:14:25 2021
Response via : Initial Calibration

