

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081821\
 Data File : VW019777.D
 Acq On : 18 Aug 2021 16:28
 Operator : SY/VA
 Sample : M3458-07
 Misc : 4.91g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AR8

Quant Time: Aug 19 05:17:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Aug 19 05:14:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	398575	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	261595	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	66596	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	163909	26.542	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 106.160%	
7) Chloroethane-d5	2.898	69	147660	27.241	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 108.960%	
11) 1,1-Dichloroethene-d2	4.025	63	174580	16.292	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 65.160%	
21) 2-Butanone-d5	7.092	46	14460	11.528	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	= 23.060%	
24) Chloroform-d	7.653	84	252007	22.990	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 91.960%	
26) 1,2-Dichloroethane-d4	8.311	65	121360	20.545	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 82.160%	
32) Benzene-d6	8.281	84	314329	21.074	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 84.280%	
36) 1,2-Dichloropropane-d6	9.275	67	145251	32.158	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 128.640%#	
41) Toluene-d8	10.329	98	238210	17.427	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 69.720%	
43) trans-1,3-Dichloroprop...	10.579	79	31395	16.825	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 67.280%	
47) 2-Hexanone-d5	10.927	63	8352	13.273	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 26.540%	
56) 1,1,2,2-Tetrachloroeth...	12.695	84	67723	18.422	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 73.680%	
66) 1,2-Dichlorobenzene-d4	13.853	152	50121	21.941	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 87.760%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081821\
Data File : VW019777.D
Acq On : 18 Aug 2021 16:28
Operator : SY/VA
Sample : M3458-07
Misc : 4.91g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AR8

Quant Time: Aug 19 05:17:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Aug 19 05:14:25 2021
Response via : Initial Calibration

