

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081918\
 Data File : VW004749.D
 Acq On : 17 Aug 2018 22:47
 Operator : SY/AP
 Sample : J4282-07
 Misc : 6.98G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SU-01-081518-B

Quant Time: Aug 18 03:09:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 18 01:49:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	582865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	967367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	853350	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	362087	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	318003	56.43	ug/l	0.00
Spiked Amount			Recovery	=	112.86%	
35) Dibromofluoromethane	7.88	113	301440	48.75	ug/l	0.00
Spiked Amount			Recovery	=	97.50%	
50) Toluene-d8	10.33	98	1055523	43.73	ug/l	0.00
Spiked Amount			Recovery	=	87.46%	
62) 4-Bromofluorobenzene	12.62	95	319598	36.81	ug/l	0.00
Spiked Amount			Recovery	=	73.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.91	84	35546	1.20	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081918\
Data File : VW004749.D
Acq On : 17 Aug 2018 22:47
Operator : SY/AP
Sample : J4282-07
Misc : 6.98G/5ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SU-01-081518-B

Quant Time: Aug 18 03:09:59 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 18 01:49:53 2018
Response via : Initial Calibration

