


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081918\
 Data File : VW004734.D
 Acq On : 17 Aug 2018 14:51
 Operator : SY/AP
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
 Title : SW846 8260
 Last Update : Sat Aug 18 01:49:53 2018



AutoFind: Scans 1810, 1811, 1812; Background Corrected with Scan 1801

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	49858	PASS
75	95	30	60	47.0	141685	PASS
95	95	100	100	100.0	301546	PASS
96	95	5	9	6.8	20584	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.0	265301	PASS
175	174	5	9	7.2	19203	PASS
176	174	95	101	96.0	254805	PASS
177	176	5	9	6.4	16394	PASS