


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW081919\
Data File : VW011995.D
Acq On : 19 Aug 2019 16:49
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Title : SW846 8260
Last Update : Sat Jul 27 00:23:30 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1743

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	32725	PASS
75	95	30	60	52.4	79992	PASS
95	95	100	100	100.0	152626	PASS
96	95	5	9	6.8	10453	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	78.2	119410	PASS
175	174	5	9	7.8	9352	PASS
176	174	95	101	96.2	114909	PASS
177	176	5	9	6.3	7286	PASS