

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081921\
 Data File : VW019801.D
 Acq On : 19 Aug 2021 12:14
 Operator : SY/VA
 Sample : M3458-08RE
 Misc : 5.43g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 COASORE

Quant Time: Aug 20 00:44:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 20 00:38:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	53957	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	50546	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	17039	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	31177	37.293	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	149.160%
7) Chloroethane-d5	2.891	69	27220	37.095	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	148.360%
11) 1,1-Dichloroethene-d2	4.025	63	33525	23.110	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	92.440%
21) 2-Butanone-d5	7.092	46	7373	43.419	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.840%
24) Chloroform-d	7.652	84	43471	29.295	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	117.200%
26) 1,2-Dichloroethane-d4	8.311	65	23163	28.966	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	115.880%
32) Benzene-d6	8.280	84	77759	26.981	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	107.920%
36) 1,2-Dichloropropane-d6	9.274	67	25173	28.843	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	115.360%
41) Toluene-d8	10.329	98	59530	22.540	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.160%
43) trans-1,3-Dichloroprop...	10.579	79	8257	22.901	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.600%
47) 2-Hexanone-d5	10.926	63	4251	34.964	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.920%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	18207	25.632	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.520%
66) 1,2-Dichlorobenzene-d4	13.853	152	15263	26.114	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.440%
Target Compounds						Qvalue
16) Methylene chloride	4.927	84	1407	1.519	ug/L #	76
42) Toluene	10.396	91	990	0.301	ug/L #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081921\
Data File : VW019801.D
Acq On : 19 Aug 2021 12:14
Operator : SY/VA
Sample : M3458-08RE
Misc : 5.43g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AS0RE

Quant Time: Aug 20 00:44:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 20 00:38:38 2021
Response via : Initial Calibration

