

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081921\
 Data File : VW019800.D
 Acq On : 19 Aug 2021 11:50
 Operator : SY/VA
 Sample : M3458-07
 Misc : 6.40g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AR8

Quant Time: Aug 20 00:44:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 20 00:38:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	420080	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	314589	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	84243	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	177922	27.336	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	109.360%
7) Chloroethane-d5	2.898	69	153653	26.896	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.600%
11) 1,1-Dichloroethene-d2	4.025	63	187694	16.619	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.480%
21) 2-Butanone-d5	7.086	46	21856	16.532	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	33.060%
24) Chloroform-d	7.653	84	262274	22.702	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.800%
26) 1,2-Dichloroethane-d4	8.311	65	135880	21.825	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.320%
32) Benzene-d6	8.281	84	414382	23.102	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	9.275	67	153686	28.293	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	113.160%
41) Toluene-d8	10.329	98	335978	20.440	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.760%
43) trans-1,3-Dichloroprop...	10.579	79	37493	16.708	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.840%
47) 2-Hexanone-d5	10.927	63	13905	18.376	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	36.760%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	85465	19.332	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	77.320%
66) 1,2-Dichlorobenzene-d4	13.853	152	65694	22.734	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW081921\
Data File : VW019800.D
Acq On : 19 Aug 2021 11:50
Operator : SY/VA
Sample : M3458-07
Misc : 6.40g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AR8

Quant Time: Aug 20 00:44:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 20 00:38:38 2021
Response via : Initial Calibration

