

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082018\
 Data File : VW004788.D
 Acq On : 20 Aug 2018 14:02
 Operator : SY/AP
 Sample : J4537-01
 Misc : 6.78G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 G.L.-WC-UST-01

Quant Time: Aug 20 14:47:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 18 01:49:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	581519	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	914379	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	678441	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	245142	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	280029	49.81	ug/l	0.00
Spiked Amount			Recovery	=	99.62%	
35) Dibromofluoromethane	7.89	113	287627	49.22	ug/l	0.00
Spiked Amount			Recovery	=	98.44%	
50) Toluene-d8	10.33	98	565851	24.80	ug/l	0.00
Spiked Amount			Recovery	=	49.60%	
62) 4-Bromofluorobenzene	12.62	95	268074	32.66	ug/l	0.00
Spiked Amount			Recovery	=	65.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.92	84	38732	1.75	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082018\
Data File : VW004788.D
Acq On : 20 Aug 2018 14:02
Operator : SY/AP
Sample : J4537-01
Misc : 6.78G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
G.L.-WC-UST-01

Quant Time: Aug 20 14:47:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 18 01:49:53 2018
Response via : Initial Calibration

