

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW082018\
 Data File : VW004805.D
 Acq On : 20 Aug 2018 21:19
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.05G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 21 01:54:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 18 01:49:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.331	0.304	8.2	91	0.00
3 P	Chloromethane	0.433	0.387	10.6	94	0.00
4 C	Vinyl Chloride	0.637	0.663	-4.1#	105	0.00
5 T	Bromomethane	0.394	0.412	-4.6	106	0.00
6 T	Chloroethane	0.385	0.410	-6.5	106	0.00
7 T	Trichlorofluoromethane	0.300	0.347	-15.7	116	0.00
8 T	Diethyl Ether	0.271	0.288	-6.3	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.542	-4.8	103	0.00
10 T	Methyl Iodide	0.785	0.851	-8.4	105	0.00
11 T	Tert butyl alcohol	0.027	0.028	-3.7	101	0.03
12 CM	1,1-Dichloroethene	0.515	0.535	-3.9#	102	0.00
13 T	Acrolein	0.036	0.035	2.8	98	0.00
14 T	Allyl chloride	0.782	0.848	-8.4	103	0.00
15 T	Acrylonitrile	0.114	0.123	-7.9	101	0.01
16 T	Acetone	0.123	0.121	1.6	97	0.01
17 T	Carbon Disulfide	1.693	1.770	-4.5	103	0.00
18 T	Methyl Acetate	0.264	0.292	-10.6	103	0.00
19 T	Methyl tert-butyl Ether	0.669	0.765	-14.3	103	0.00
20 T	Methylene Chloride	0.673	0.616	8.5	106	0.00
21 T	trans-1,2-Dichloroethene	0.548	0.584	-6.6	101	0.00
22 T	Diisopropyl ether	1.503	1.729	-15.0	106	0.00
23 T	Vinyl Acetate	0.850	0.962	-13.2	101	0.00
24 P	1,1-Dichloroethane	0.977	1.049	-7.4	103	0.00
25 T	2-Butanone	0.176	0.183	-4.0	98	0.00
26 T	2,2-Dichloropropane	0.564	0.584	-3.5	103	0.00
27 T	cis-1,2-Dichloroethene	0.590	0.641	-8.6	102	0.00
28 T	Bromochloromethane	0.411	0.426	-3.6	94	0.00
29 T	Tetrahydrofuran	0.090	0.098	-8.9	99	0.00
30 C	Chloroform	0.961	1.019	-6.0#	102	0.00
31 T	Cyclohexane	0.945	0.976	-3.3	101	0.00
32 T	1,1,1-Trichloroethane	0.761	0.818	-7.5	103	0.00
33 S	1,2-Dichloroethane-d4	0.483	0.482	0.2	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.320	0.321	-0.3	95	0.00
36 T	1,1-Dichloropropene	0.515	0.532	-3.3	101	0.00
37 T	Ethyl Acetate	0.205	0.215	-4.9	98	0.00
38 T	Carbon Tetrachloride	0.474	0.486	-2.5	100	0.00
39 T	Methylcyclohexane	0.644	0.692	-7.5	100	0.00
40 TM	Benzene	1.497	1.556	-3.9	101	0.00
41 T	Methacrylonitrile	0.117	0.140	-19.7	109	0.00
42 TM	1,2-Dichloroethane	0.384	0.402	-4.7	102	0.00
43 T	Isopropyl Acetate	0.369	0.402	-8.9	104	0.00
44 TM	Trichloroethene	0.380	0.392	-3.2	101	0.00
45 C	1,2-Dichloropropane	0.366	0.383	-4.6#	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW082018\
 Data File : VW004805.D
 Acq On : 20 Aug 2018 21:19
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.05G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 21 01:54:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 18 01:49:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.181	0.187	-3.3	101	0.00
47 T	Bromodichloromethane	0.433	0.457	-5.5	102	0.00
48 T	Methyl methacrylate	0.173	0.191	-10.4	103	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	107	0.01
50 S	Toluene-d8	1.247	1.255	-0.6	94	0.00
51 T	4-Methyl-2-Pentanone	0.249	0.273	-9.6	102	0.00
52 CM	Toluene	0.898	0.955	-6.3#	103	0.00
53 T	t-1,3-Dichloropropene	0.436	0.470	-7.8	103	0.00
54 T	cis-1,3-Dichloropropene	0.520	0.557	-7.1	103	0.00
55 T	1,1,2-Trichloroethane	0.258	0.270	-4.7	102	0.00
56 T	Ethyl methacrylate	0.305	0.346	-13.4	103	0.00
57 T	1,3-Dichloropropane	0.445	0.472	-6.1	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.142	0.166	-16.9	108	0.00
59 T	2-Hexanone	0.169	0.183	-8.3	100	0.00
60 T	Dibromochloromethane	0.289	0.308	-6.6	102	0.00
61 T	1,2-Dibromoethane	0.246	0.260	-5.7	103	0.00
62 S	4-Bromofluorobenzene	0.449	0.455	-1.3	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.342	0.358	-4.7	103	0.00
65 PM	Chlorobenzene	1.076	1.123	-4.4	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.381	-5.0	103	0.00
67 C	Ethyl Benzene	1.824	1.978	-8.4#	102	0.00
68 T	m/p-Xylenes	0.710	0.768	-8.2	102	0.00
69 T	o-Xylene	0.657	0.726	-10.5	104	0.00
70 T	Styrene	1.104	1.218	-10.3	102	0.00
71 P	Bromoform	0.192	0.203	-5.7	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.402	3.811	-12.0	102	0.00
74 T	N-amyl acetate	0.713	0.792	-11.1	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.630	0.675	-7.1	102	0.00
76 T	1,2,3-Trichloropropane	0.441	0.442	-0.2	101	0.00
77 T	Bromobenzene	0.823	0.886	-7.7	102	0.00
78 T	n-propylbenzene	4.151	4.630	-11.5	101	0.00
79 T	2-Chlorotoluene	2.393	2.646	-10.6	103	0.00
80 T	1,3,5-Trimethylbenzene	2.890	3.236	-12.0	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.190	0.204	-7.4	99	0.00
82 T	4-Chlorotoluene	2.542	2.769	-8.9	101	0.00
83 T	tert-Butylbenzene	2.472	2.779	-12.4	102	0.00
84 T	1,2,4-Trimethylbenzene	2.973	3.322	-11.7	101	0.00
85 T	sec-Butylbenzene	3.690	4.105	-11.2	101	0.00
86 T	p-Isopropyltoluene	3.211	3.599	-12.1	101	0.00
87 T	1,3-Dichlorobenzene	1.675	1.795	-7.2	102	0.00
88 T	1,4-Dichlorobenzene	1.674	1.755	-4.8	100	0.00
89 T	n-Butylbenzene	3.105	3.398	-9.4	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW082018\
Data File : VW004805.D
Acq On : 20 Aug 2018 21:19
Operator : SY/AP
Sample : VSTDCCC050
Misc : 5.05G/5ML/MSVOA W/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Aug 21 01:54:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 18 01:49:53 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.585	0.631	-7.9	103	0.00
91 T	1,2-Dichlorobenzene	1.504	1.601	-6.4	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.098	0.105	-7.1	101	0.00
93 T	1,2,4-Trichlorobenzene	1.064	1.144	-7.5	100	0.00
94 T	Hexachlorobutadiene	0.653	0.684	-4.7	100	0.00
95 T	Naphthalene	1.840	2.084	-13.3	102	0.00
96 T	1,2,3-Trichlorobenzene	0.943	1.019	-8.1	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6