

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082019\
 Data File : VW012024.D
 Acq On : 20 Aug 2019 12:01
 Operator : SY/VA
 Sample : K4391-06
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 N-WALL-0

Quant Time: Aug 20 17:35:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	261893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	436743	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	351627	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	152653	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	160585	53.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.80%	
35) Dibromofluoromethane	7.88	113	131195	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
50) Toluene-d8	10.32	98	536565	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
62) 4-Bromofluorobenzene	12.62	95	236059	66.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	132.00%	
Target Compounds						
16) Acetone	4.13	43	42084	43.993	ug/l	99
20) Methylene Chloride	4.90	84	5139	1.558	ug/l	86
78) n-propylbenzene	12.80	91	27611	1.992	ug/l	89
80) 1,3,5-Trimethylbenzene	12.94	105	19597	2.053	ug/l	98
83) tert-Butylbenzene	13.21	119	39425	4.939	ug/l	95
84) 1,2,4-Trimethylbenzene	13.25	105	20718	2.154	ug/l	96
85) sec-Butylbenzene	13.38	105	76544	6.720	ug/l #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082019\
Data File : VW012024.D
Acq On : 20 Aug 2019 12:01
Operator : SY/VA
Sample : K4391-06
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
N-WALL-0

Quant Time: Aug 20 17:35:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:23:30 2019
Response via : Initial Calibration

