

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082019\
 Data File : VW012035.D
 Acq On : 20 Aug 2019 17:28
 Operator : SY/VA
 Sample : VW0820SBL02
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0820SBL02

Quant Time: Aug 21 01:53:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	250475	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	415305	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	350454	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	148901	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	155242	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
35) Dibromofluoromethane	7.88	113	127096	54.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.10%	
50) Toluene-d8	10.32	98	507253	53.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.34%	
62) 4-Bromofluorobenzene	12.62	95	168843	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082019\
Data File : VW012035.D
Acq On : 20 Aug 2019 17:28
Operator : SY/VA
Sample : VW0820SBL02
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0820SBL02

Quant Time: Aug 21 01:53:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:23:30 2019
Response via : Initial Calibration

