

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082019\
 Data File : VW012038.D
 Acq On : 20 Aug 2019 19:12
 Operator : SY/VA
 Sample : K4388-17
 Misc : 3.36G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T6-E1-(0)

Quant Time: Aug 21 02:00:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	72703	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	130596	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	107758	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	29379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	89051	106.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	213.34%	
35) Dibromofluoromethane	7.88	113	57338	78.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	156.52%	
50) Toluene-d8	10.32	98	144349	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
62) 4-Bromofluorobenzene	12.62	95	40235	37.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.24%	
Target Compounds						
16) Acetone	4.13	43	4921	18.531	ug/l	98
20) Methylene Chloride	4.92	84	9057	9.889	ug/l	94
29) Tetrahydrofuran	7.53	42	4844	24.387	ug/l	96
44) Trichloroethene	9.09	130	1833	2.005	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082019\
Data File : VW012038.D
Acq On : 20 Aug 2019 19:12
Operator : SY/VA
Sample : K4388-17
Misc : 3.36G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T6-E1-(0)

Quant Time: Aug 21 02:00:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:23:30 2019
Response via : Initial Calibration

