

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082019\
 Data File : VW012034.D
 Acq On : 20 Aug 2019 16:27
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 21 01:51:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.279	0.299	-7.2	106	0.00
3 P	Chloromethane	0.483	0.489	-1.2	106	0.00
4 C	Vinyl Chloride	0.582	0.621	-6.7#	104	0.00
5 T	Bromomethane	0.322	0.324	-0.6	101	0.00
6 T	Chloroethane	0.332	0.361	-8.7	105	0.00
7 T	Trichlorofluoromethane	0.260	0.319	-22.7#	116	0.00
8 T	Diethyl Ether	0.295	0.308	-4.4	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.493	0.555	-12.6	108	0.00
10 T	Methyl Iodide	0.697	0.764	-9.6	103	0.00
11 T	Tert butyl alcohol	0.048	0.042	12.5	99	-0.01
12 CM	1,1-Dichloroethene	0.521	0.573	-10.0#	106	0.00
13 T	Acrolein	0.041	0.039	4.9	99	0.00
14 T	Allyl chloride	1.108	1.203	-8.6	104	0.00
15 T	Acrylonitrile	0.147	0.156	-6.1	101	0.00
16 T	Acetone	0.183	0.173	5.5	91	0.00
17 T	Carbon Disulfide	1.630	1.647	-1.0	96	0.00
18 T	Methyl Acetate	0.410	0.414	-1.0	96	0.00
19 T	Methyl tert-butyl Ether	0.826	0.895	-8.4	101	0.00
20 T	Methylene Chloride	0.630	0.615	2.4	101	0.00
21 T	trans-1,2-Dichloroethene	0.561	0.595	-6.1	103	0.00
22 T	Diisopropyl ether	2.088	2.206	-5.7	101	0.00
23 T	Vinyl Acetate	1.312	1.405	-7.1	100	0.00
24 P	1,1-Dichloroethane	1.118	1.201	-7.4	103	0.00
25 T	2-Butanone	0.239	0.234	2.1	91	0.00
26 T	2,2-Dichloropropane	0.688	0.683	0.7	103	0.00
27 T	cis-1,2-Dichloroethene	0.618	0.643	-4.0	101	0.00
28 T	Bromochloromethane	0.496	0.539	-8.7	115	0.00
29 T	Tetrahydrofuran	0.137	0.144	-5.1	99	0.00
30 C	Chloroform	1.032	1.087	-5.3#	103	0.00
31 T	Cyclohexane	1.073	1.102	-2.7	102	0.00
32 T	1,1,1-Trichloroethane	0.786	0.868	-10.4	106	0.00
33 S	1,2-Dichloroethane-d4	0.574	0.553	3.7	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.281	0.293	-4.3	100	0.00
36 T	1,1-Dichloropropene	0.520	0.580	-11.5	104	0.00
37 T	Ethyl Acetate	0.299	0.313	-4.7	96	0.00
38 T	Carbon Tetrachloride	0.445	0.499	-12.1	104	0.00
39 T	Methylcyclohexane	0.572	0.657	-14.9	106	0.00
40 TM	Benzene	1.456	1.565	-7.5	102	0.00
41 T	Methacrylonitrile	0.179	0.195	-8.9	102	0.00
42 TM	1,2-Dichloroethane	0.456	0.477	-4.6	98	0.00
43 T	Isopropyl Acetate	0.550	0.600	-9.1	100	0.00
44 TM	Trichloroethene	0.350	0.389	-11.1	104	0.00
45 C	1,2-Dichloropropane	0.390	0.423	-8.5#	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082019\
 Data File : VW012034.D
 Acq On : 20 Aug 2019 16:27
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 21 01:51:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.178	0.189	-6.2	99	0.00
47 T	Bromodichloromethane	0.467	0.510	-9.2	102	0.00
48 T	Methyl methacrylate	0.264	0.292	-10.6	99	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	102	0.00
50 S	Toluene-d8	1.148	1.195	-4.1	100	0.00
51 T	4-Methyl-2-Pentanone	0.286	0.312	-9.1	98	0.00
52 CM	Toluene	0.879	0.966	-9.9#	103	0.00
53 T	t-1,3-Dichloropropene	0.487	0.531	-9.0	99	0.00
54 T	cis-1,3-Dichloropropene	0.579	0.638	-10.2	100	0.00
55 T	1,1,2-Trichloroethane	0.252	0.270	-7.1	99	0.00
56 T	Ethyl methacrylate	0.369	0.417	-13.0	100	0.00
57 T	1,3-Dichloropropane	0.477	0.516	-8.2	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.168	0.210	-25.0#	112	0.00
59 T	2-Hexanone	0.202	0.224	-10.9	97	0.00
60 T	Dibromochloromethane	0.280	0.309	-10.4	101	0.00
61 T	1,2-Dibromoethane	0.239	0.257	-7.5	99	0.00
62 S	4-Bromofluorobenzene	0.409	0.428	-4.6	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.326	0.368	-12.9	105	0.00
65 PM	Chlorobenzene	1.005	1.121	-11.5	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.350	0.392	-12.0	100	0.00
67 C	Ethyl Benzene	1.915	2.213	-15.6#	104	0.00
68 T	m/p-Xylenes	0.696	0.791	-13.6	103	0.00
69 T	o-Xylene	0.647	0.739	-14.2	102	0.00
70 T	Styrene	1.119	1.285	-14.8	102	0.00
71 P	Bromoform	0.179	0.204	-14.0	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.730	4.256	-14.1	105	0.00
74 T	N-amyl acetate	1.168	1.366	-17.0	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.702	0.777	-10.7	100	0.00
76 T	1,2,3-Trichloropropane	0.529	0.532	-0.6	98	0.00
77 T	Bromobenzene	0.824	0.924	-12.1	102	0.00
78 T	n-propylbenzene	4.540	5.220	-15.0	105	0.00
79 T	2-Chlorotoluene	2.615	2.922	-11.7	104	0.00
80 T	1,3,5-Trimethylbenzene	3.127	3.562	-13.9	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.230	0.263	-14.3	100	0.00
82 T	4-Chlorotoluene	2.744	3.078	-12.2	103	0.00
83 T	tert-Butylbenzene	2.615	3.075	-17.6	107	0.00
84 T	1,2,4-Trimethylbenzene	3.150	3.582	-13.7	104	0.00
85 T	sec-Butylbenzene	3.731	4.392	-17.7	107	0.00
86 T	p-Isopropyltoluene	3.385	3.959	-17.0	107	0.00
87 T	1,3-Dichlorobenzene	1.617	1.804	-11.6	103	0.00
88 T	1,4-Dichlorobenzene	1.608	1.773	-10.3	103	0.00
89 T	n-Butylbenzene	3.378	3.992	-18.2	108	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082019\
Data File : VW012034.D
Acq On : 20 Aug 2019 16:27
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Aug 21 01:51:04 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:23:30 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.599	0.689	-15.0	104	0.00
91 T	1,2-Dichlorobenzene	1.430	1.612	-12.7	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.129	-7.5	95	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.163	-15.4	104	0.00
94 T	Hexachlorobutadiene	0.628	0.752	-19.7	110	0.00
95 T	Naphthalene	1.787	2.120	-18.6	100	0.00
96 T	1,2,3-Trichlorobenzene	0.883	1.011	-14.5	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6