

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082118\
 Data File : VW004873.D
 Acq On : 22 Aug 2018 07:02
 Operator : SY/AP
 Sample : J4271-13
 Misc : 6.97G/5ML/MSVOA W/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 HR-06-082018-C

Quant Time: Aug 22 08:32:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 18 01:49:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	547049	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	884528	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	697535	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	273336	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	301321	56.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.94%	
35) Dibromofluoromethane	7.89	113	300498	53.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.30%	
50) Toluene-d8	10.33	98	771323	34.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.90%	
62) 4-Bromofluorobenzene	12.62	95	290357	36.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.14%	
Target Compounds						
16) Acetone	4.14	43	40261	29.82	ug/l	100
52) Toluene	10.40	92	37898	2.39	ug/l	100
68) m/p-Xylenes	11.84	106	10389	1.05	ug/l	98
69) o-Xylene	12.17	106	10799	1.18	ug/l	99
80) 1,3,5-Trimethylbenzene	12.95	105	20572	1.30	ug/l	99
95) Naphthalene	15.38	128	11982	4.50	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082118\
Data File : VW004873.D
Acq On : 22 Aug 2018 07:02
Operator : SY/AP
Sample : J4271-13
Misc : 6.97G/5ML/MSVOA W/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
HR-06-082018-C

Quant Time: Aug 22 08:32:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 18 01:49:53 2018
Response via : Initial Calibration

