


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW082118\
 Data File : VW004858.D
 Acq On : 21 Aug 2018 23:45
 Operator : SY/AP
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081918S.M
 Title : SW846 8260
 Last Update : Sat Aug 18 01:49:53 2018



AutoFind: Scans 1810, 1811, 1812; Background Corrected with Scan 1801

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	48309	PASS
75	95	30	60	46.7	136778	PASS
95	95	100	100	100.0	292714	PASS
96	95	5	9	6.6	19319	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	84.7	248000	PASS
175	174	5	9	7.3	18037	PASS
176	174	95	101	96.5	239381	PASS
177	176	5	9	6.8	16366	PASS