

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082119\
 Data File : VW012089.D
 Acq On : 21 Aug 2019 23:45
 Operator : SY/VA
 Sample : K4403-11
 Misc : 5.23G/5ML/MSVOA W/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T1-G-(0)

Quant Time: Aug 22 07:08:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	238679	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	403489	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	322551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	113717	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	154322	56.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.62%	
35) Dibromofluoromethane	7.88	113	122990	54.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.66%	
50) Toluene-d8	10.32	98	486199	52.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.92%	
62) 4-Bromofluorobenzene	12.62	95	143482	43.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.84%	
Target Compounds						
						Qvalue
8) Diethyl Ether	3.68	74	3818	2.713	ug/l	83
16) Acetone	4.14	43	9814	11.257	ug/l	96
20) Methylene Chloride	4.92	84	34470	11.464	ug/l	98
29) Tetrahydrofuran	7.54	42	5925	9.086	ug/l	98
95) Naphthalene	15.37	128	3419	2.796	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082119\
 Data File : VW012089.D
 Acq On : 21 Aug 2019 23:45
 Operator : SY/VA
 Sample : K4403-11
 Misc : 5.23G/5ML/MSVOA W/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T1-G-(0)

Quant Time: Aug 22 07:08:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

