

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082220\
 Data File : VW016309.D
 Acq On : 21 Aug 2020 16:37
 Operator : SY/VA
 Sample : L3734-04
 Misc : 6.13G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EROC-BF-1C

Quant Time: Aug 22 06:44:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081420S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 17 08:01:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	177357	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	302726	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	245246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	88437	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	70923	41.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.02%	
35) Dibromofluoromethane	7.88	113	32815	17.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	35.14%	
50) Toluene-d8	10.33	98	356493	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	12.62	95	104945	40.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.22%	
Target Compounds						
8) Diethyl Ether	3.69	74	2515	2.754	ug/l	99
16) Acetone	4.12	43	7720	23.300	ug/l	91
20) Methylene Chloride	4.92	84	46658	19.672	ug/l	89
84) 1,2,4-Trimethylbenzene	13.26	105	9311	1.626	ug/l	96
95) Naphthalene	15.37	128	8712	6.063	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082220\
Data File : VW016309.D
Acq On : 21 Aug 2020 16:37
Operator : SY/VA
Sample : L3734-04
Misc : 6.13G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EROC-BF-1C

Quant Time: Aug 22 06:44:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W081420S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 17 08:01:41 2020
Response via : Initial Calibration

