

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082223\
 Data File : VW026805.D
 Acq On : 22 Aug 2023 17:02
 Operator : SY/MD
 Sample : 04108-09
 Misc : 5.93g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCHX0

Quant Time: Aug 23 01:32:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Aug 23 01:29:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	497767	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	502614	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	246963	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	102089	12.476	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	49.920%		
7) Chloroethane-d5	2.893	69	89500	14.709	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	58.840%		
11) 1,1-Dichloroethene-d2	4.021	65	60529	15.040	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	60.160%		
21) 2-Butanone-d5	7.081	46	104481	44.425	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	88.840%		
24) Chloroform-d	7.648	84	337469	20.200	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	80.800%		
26) 1,2-Dichloroethane-d4	8.307	65	202088	21.920	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	87.680%		
32) Benzene-d6	8.270	84	644795	18.614	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	74.440%		
36) 1,2-Dichloropropane-d6	9.270	67	209146	19.383	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	77.520%		
41) Toluene-d8	10.325	98	562478	18.916	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	75.680%		
43) trans-1,3-Dichloroprop...	10.581	79	87688	19.683	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	78.720%		
47) 2-Hexanone-d5	10.922	63	53379	47.081	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	94.160%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	203420	23.872	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	95.480%		
66) 1,2-Dichlorobenzene-d4	13.848	152	210545	23.819	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	95.280%		
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	26983	2.802	ug/L	94
46) Tetrachloroethene	10.861	164	34832	5.143	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082223\
Data File : VW026805.D
Acq On : 22 Aug 2023 17:02
Operator : SY/MD
Sample : 04108-09
Misc : 5.93g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCHX0

Quant Time: Aug 23 01:32:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 23 01:29:10 2023
Response via : Initial Calibration

