

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082319\
 Data File : VW012141.D
 Acq On : 23 Aug 2019 18:58
 Operator : SY/VA
 Sample : K4474-03
 Misc : 5.69G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T1-2-F2-(0)

Quant Time: Aug 26 07:52:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	224830	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	377416	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	307893	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	113240	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	147672	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
35) Dibromofluoromethane	7.88	113	116833	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
50) Toluene-d8	10.32	98	461076	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
62) 4-Bromofluorobenzene	12.62	95	139660	43.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.74%	
Target Compounds						
20) Methylene Chloride	4.92	84	19414	7.517	ug/l	96
29) Tetrahydrofuran	7.53	42	4819	8.652	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082319\
Data File : VW012141.D
Acq On : 23 Aug 2019 18:58
Operator : SY/VA
Sample : K4474-03
Misc : 5.69G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T1-2-F2-(0)

Quant Time: Aug 26 07:52:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

