

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082319\
 Data File : VW012147.D
 Acq On : 23 Aug 2019 21:33
 Operator : SY/VA
 Sample : K4474-10
 Misc : 6.01G/5ML/MSVOA W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T1-2-K1-(0)

Quant Time: Aug 26 08:08:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	213583	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	360041	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	292481	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	107603	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	143105	56.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.72%	
35) Dibromofluoromethane	7.88	113	112933	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
50) Toluene-d8	10.32	98	439632	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	12.62	95	132181	43.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.06%	
Target Compounds						
20) Methylene Chloride	4.92	84	19640	8.005	ug/l	97
29) Tetrahydrofuran	7.54	42	5792	10.946	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082319\
Data File : VW012147.D
Acq On : 23 Aug 2019 21:33
Operator : SY/VA
Sample : K4474-10
Misc : 6.01G/5ML/MSVOA W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T1-2-K1-(0)

Quant Time: Aug 26 08:08:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

